


7470A

LB 90690
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34852



A= 0.0000e+000

B= 2.8735e-004

C= -2.0459e-002

Rho= 0.9997394

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	192	0.000	192				
0.2	0.200	0.200	0.000	768	0.0 %	768				
2.5	2.500	2.539	0.039	8906	0.0 %	8906				
5.0	5.000	4.829	-0.171	16877	0.0 %	16877				
7.5	7.500	7.603	0.103	26531	0.0 %	26531				
10.0	10.000	9.994	-0.006	34852	0.0 %	34852				

LB90690 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Oct 2017 10:41:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	192	-	0.0000	7470A	PPB	10 Oct 2017 10:43:39	S	Std
0.2 - 1	80.2	768	-	0.2000	7470A	PPB	10 Oct 2017 10:45:48	S	Std
2.5 - 1	82.5	8906	-	2.5000	7470A	PPB	10 Oct 2017 10:47:56	S	Std
5.0 - 1	55	16877	-	5.0000	7470A	PPB	10 Oct 2017 10:50:04	S	Std
7.5 - 1	57.5	26531	-	7.5000	7470A	PPB	10 Oct 2017 10:52:12	S	Std
10.0 - 1	510	34852	-	10.0000	7470A	PPB	10 Oct 2017 10:54:20	S	Std
ICV99 - 1	ICV99	13822	3.9513	-	7470A	PPB	10 Oct 2017 10:57:14	U	SMPL
ICB99 - 1	ICB99	42	-0.0084	-	7470A	PPB	10 Oct 2017 10:59:21	U	SMPL
CCV62 - 1	CCV62	16960	4.8529	-	7470A	PPB	10 Oct 2017 11:01:27	U	SMPL
CCB62 - 1	CCB62	48	-0.0067	-	7470A	PPB	10 Oct 2017 11:03:34	U	SMPL
CRA - 1	CRA	787	0.2057	-	7470A	PPB	10 Oct 2017 11:05:41	U	SMPL
HighStd - 1	HighStd	35260	10.1114	-	7470A	PPB	10 Oct 2017 11:07:48	U	SMPL
ChkStd - 1	ChkStd	24678	7.0707	-	7470A	PPB	10 Oct 2017 11:09:54	U	SMPL
PB103105BL - 1	PBW	-259	-0.0949	-	7470A	PPB	10 Oct 2017 11:12:03	U	SMPL
PB103105BS - 1	LCSW	12811	3.6607	-	7470A	PPB	10 Oct 2017 11:14:12	U	SMPL
PB103105TB - 1	PB103023TB	-168	-0.0687	-	7470A	PPB	10 Oct 2017 11:16:20	U	SMPL
I5745-03 - 1	HR-06-100617-A	-82	-0.0440	-	7470A	PPB	10 Oct 2017 11:18:28	U	SMPL
I5745-06 - 1	HR-06-100617-B	-39	-0.0317	-	7470A	PPB	10 Oct 2017 11:20:36	U	SMPL
I5745-09 - 1	HR-06-100617-C	-33	-0.0299	-	7470A	PPB	10 Oct 2017 11:22:44	U	SMPL
I5748-03 - 1	CL-01-100617-A	-4	-0.0216	-	7470A	PPB	10 Oct 2017 11:24:51	U	SMPL
CCV63 - 1	CCV63	17183	4.9170	-	7470A	PPB	10 Oct 2017 11:26:58	U	SMPL
CCB63 - 1	CCB63	65	-0.0018	-	7470A	PPB	10 Oct 2017 11:29:05	U	SMPL
I5748-06 - 1	CL-01-100617-B	-23	-0.0271	-	7470A	PPB	10 Oct 2017 11:31:12	U	SMPL
I5748-09 - 1	CL-01-100617-C	0	-0.0205	-	7470A	PPB	10 Oct 2017 11:33:18	U	SMPL
I5749-02 - 1	SP-COMP	-7	-0.0225	-	7470A	PPB	10 Oct 2017 11:35:25	U	SMPL
I5749-02DUP - 1	SP-COMPDUP	-51	-0.0351	-	7470A	PPB	10 Oct 2017 11:37:32	U	SMPL
I5749-02MS - 1	SP-COMPMS	13058	3.7317	-	7470A	PPB	10 Oct 2017 11:39:39	U	SMPL
I5749-02MSD - 1	SP-COMPMSD	13003	3.7159	-	7470A	PPB	10 Oct 2017 11:41:47	U	SMPL
I5749-02LX5 - 1		4	-0.0193	-	7470A	PPB	10 Oct 2017 11:43:55	U	SMPL
I5749-02A - 1		11469	3.2751	-	7470A	PPB	10 Oct 2017 11:46:03	U	SMPL
CCV64 - 1	CCV64	17054	4.8800	-	7470A	PPB	10 Oct 2017 11:48:12	U	SMPL
CCB64 - 1	CCB64	19	-0.0150	-	7470A	PPB	10 Oct 2017 11:50:19	U	SMPL