


ISM02.4

LB 90693
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40829



A= 0.0000e+000

B= 2.4606e-004

C= -1.6362e-002

Rho= 0.9999817

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.020	0.020	149	0.000	149					-
0.2	0.200	0.190	-0.010	839	0.0 %	839					-5
2.5	2.500	2.513	0.013	10278	0.0 %	10278					-1
5.0	5.000	4.974	-0.026	20281	0.0 %	20281					-1
7.5	7.500	7.473	-0.027	30436	0.0 %	30436					0
10.0	10.000	10.030	0.030	40829	0.0 %	40829					0

NB
10/10/13

LB90693
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	149	-		0 ISM02.4	PPB	10/10/2017 12:05	Std
	0.2 S0.2	839	-		0.2 ISM02.4	PPB	10/10/2017 12:07	Std
	2.5 S2.5	10278	-		2.5 ISM02.4	PPB	10/10/2017 12:09	Std
	5 S5	20281	-		5 ISM02.4	PPB	10/10/2017 12:11	Std
	7.5 S7.5	30436	-		7.5 ISM02.4	PPB	10/10/2017 12:14	Std
	10 S10	40829	-		10 ISM02.4	PPB	10/10/2017 12:16	Std
ICV	ICV	16404	4.02 -		ISM02.4	PPB	10/10/2017 12:19	SMPL
ICB	ICB	-70	-0.0336 -		ISM02.4	PPB	10/10/2017 12:21	SMPL
CCV065	CCV065	20463	5.0188 -		ISM02.4	PPB	10/10/2017 12:24	SMPL
CCB065	CCB065	-91	-0.0388 -		ISM02.4	PPB	10/10/2017 12:26	SMPL
PB103098BL	PBW003	-73	-0.0343 -		ISM02.4	PPB	10/10/2017 12:28	SMPL
I5722-01	TWP-B-39	303	0.0582 -		ISM02.4	PPB	10/10/2017 12:30	SMPL
I5722-01D	TWP-B-39D	293	0.0557 -		ISM02.4	PPB	10/10/2017 12:33	SMPL
I5722-01S	TWP-B-39S	4644	1.1264 -		ISM02.4	PPB	10/10/2017 12:35	SMPL
I5722-02	TWP-B-59	60	-0.0016 -		ISM02.4	PPB	10/10/2017 12:37	SMPL
I5722-03	TWP-B-55	25	-0.0102 -		ISM02.4	PPB	10/10/2017 12:39	SMPL
I5722-04	TWP-B-57	162	0.0235 -		ISM02.4	PPB	10/10/2017 12:42	SMPL
I5722-05	TWP-B-32	1968	0.4679 -		ISM02.4	PPB	10/10/2017 12:44	SMPL
CCV066	CCV066	20428	5.0102 -		ISM02.4	PPB	10/10/2017 12:46	SMPL
CCB066	CCB066	-89	-0.0383 -		ISM02.4	PPB	10/10/2017 12:49	SMPL