


7471A

LB 90714
INSTRUMENT ID: CV2 μ Abs.:

43982



Linear

A= 0.0000e+000

B= 2.2800e-004

C= 2.0902e-002

Rho= 0.9999562

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	8	0.000	8				
0.2	0.200	0.229	0.029	914	0.0 %	914				
2.5	2.500	2.466	-0.034	10723	0.0 %	10723				
5.0	5.000	4.966	-0.034	21691	0.0 %	21691				
7.5	7.500	7.467	-0.033	32658	0.0 %	32658	0			
10.0	10.000	10.049	0.049	43982	0.0 %	43982				

lb90714 instrument id : cv1

Method: 7471A		Operator: Admin		Date of Analysis: 10 Oct 2017 17:44:11				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	8	-	0.00007471A	PPB	10 Oct 2017 17:47:00	S	Std
0.2 - 1	50.2	914	-	0.20007471A	PPB	10 Oct 2017 17:49:18	S	Std
2.5 - 1	52.5	10723	-	2.50007471A	PPB	10 Oct 2017 17:51:35	S	Std
5.0 - 1	55	21691	-	5.00007471A	PPB	10 Oct 2017 17:53:53	S	Std
7.5 - 1	57.5	32658	-	7.50007471A	PPB	10 Oct 2017 17:56:10	S	Std
10.0 - 1	510	43982	-	10.00007471A	PPB	10 Oct 2017 17:58:27	S	Std
ICV01 - 1	ICV01	16540	3.7920	-7471A	PPB	10 Oct 2017 18:01:11	U	SMPL
ICB01 - 1	ICB01	-264	-0.0393	-7471A	PPB	10 Oct 2017 18:03:28	U	SMPL
CCV71 - 1	CCV71	21545	4.9332	-7471A	PPB	10 Oct 2017 18:05:44	U	SMPL
CCB71 - 1	CCB71	-285	-0.0441	-7471A	PPB	10 Oct 2017 18:07:59	U	SMPL
CRA - 1	CRA	883	0.2222	-7471A	PPB	10 Oct 2017 18:10:15	U	SMPL
HighStd - 1	HighStd	43989	10.0504	-7471A	PPB	10 Oct 2017 18:12:30	U	SMPL
ChkStd - 1	ChkStd	30011	6.8634	-7471A	PPB	10 Oct 2017 18:14:46	U	SMPL
PB103120BL - 1	PBS	-172	-0.0183	-7471A	PPB	10 Oct 2017 18:17:05	U	SMPL
PB103120BS - 1	LCSS	15355	3.5219	-7471A	PPB	10 Oct 2017 18:19:23	U	SMPL
I5752-01 - 1	ENV-116-8(0-5)	8770	2.0205	-7471A	PPB	10 Oct 2017 18:21:40	U	SMPL
I5752-01DUP - 1	ENV-116-8(0-5)DUP	8608	1.9835	-7471A	PPB	10 Oct 2017 18:23:57	U	SMPL
I5752-01MS - 1	ENV-116-8(0-5)MS	23773	5.4412	-7471A	PPB	10 Oct 2017 18:26:15	U	SMPL
I5752-01MSD - 1	ENV-116-8(0-5)MSD	24102	5.5162	-7471A	PPB	10 Oct 2017 18:28:32	U	SMPL
I5752-02 - 1	ENV-116-8(5-10)	8830	2.0341	-7471A	PPB	10 Oct 2017 18:30:49	U	SMPL
CCV72 - 1	CCV72	21451	4.9117	-7471A	PPB	10 Oct 2017 18:33:05	U	SMPL
CCB72 - 1	CCB72	-234	-0.0325	-7471A	PPB	10 Oct 2017 18:35:21	U	SMPL
I5752-03 - 1	ENV-118-7(0-5)	27015	6.1803	-7471A	PPB	10 Oct 2017 18:37:37	U	SMPL
I5752-04 - 1	ENV-118-7(5-10)	33707	7.7061	-7471A	PPB	10 Oct 2017 18:39:53	U	SMPL
I5752-05 - 1	ENV-119-3(0-5)	8152	1.8796	-7471A	PPB	10 Oct 2017 18:42:11	U	SMPL
I5752-06 - 1	ENV-119-3(5-10)	100085	22.8403	-7471A	PPB	10 Oct 2017 18:44:28	U	SMPL
I5752-08 - 1	ENV-114-3(0-3)	228989	52.2305	-7471A	PPB	10 Oct 2017 18:49:05	U	SMPL
I5752-09 - 1	EPE-P-6(45-47)	1221	0.2993	-7471A	PPB	10 Oct 2017 18:54:25	U	SMPL
I5752-10 - 1	TEST-PIT-116-C(0-5)	8512	1.9616	-7471A	PPB	10 Oct 2017 18:56:42	U	SMPL
I5763-01 - 1	220-2(2TRH)-COMP	31472	7.1965	-7471A	PPB	10 Oct 2017 18:59:00	U	SMPL
I5766-04 - 1	RS-CONCRETE	7761	1.7904	-7471A	PPB	10 Oct 2017 19:01:16	U	SMPL
CCV73 - 1	CCV73	21353	4.8894	-7471A	PPB	10 Oct 2017 19:03:34	U	SMPL
CCB73 - 1	CCB73	-289	-0.0450	-7471A	PPB	10 Oct 2017 19:05:51	U	SMPL
I5752-01LX5 - 1		1635	0.3937	-7471A	PPB	10 Oct 2017 19:08:08	U	SMPL
I5752-01A - 1		25282	5.7852	-7471A	PPB	10 Oct 2017 19:10:24	U	SMPL
I5752-07 - 1	EPE-P-6(25-27)	112	0.0464	-7471A	PPB	10 Oct 2017 19:12:40	U	SMPL
I5752-06DLX5 - 1		19430	4.4510	-7471A	PPB	10 Oct 2017 19:14:58	U	SMPL
I5752-08DLX10 - 1		23736	5.4327	-7471A	PPB	10 Oct 2017 19:17:15	U	SMPL
CCV74 - 1	CCV74	21523	4.9282	-7471A	PPB	10 Oct 2017 19:19:32	U	SMPL
CCB74 - 1	CCB74	-190	-0.0224	-7471A	PPB	10 Oct 2017 19:21:50	U	SMPL