



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90716

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/10/2017	16:35
2	CAL2	1	Water			7.00	10/10/2017	16:36
3	CAL3	1	Water			10.02	10/10/2017	16:37
4	ICV	1	Water			6.99	10/10/2017	16:38
5	CCV1	1	Water			2.01	10/10/2017	16:40
6	I5775-03	1	Solid	20.02	20	7.14	10/10/2017	16:41
7	I5780-04	1	Solid	20.03	20	8.30	10/10/2017	16:43
8	I5780-08	1	Solid	20.01	20	8.40	10/10/2017	16:45
9	I5780-08DUP	1	Solid	20.03	20	8.41	10/10/2017	16:46
10	CCV2	1	Water			12.01	10/10/2017	16:47

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5580

WorkList ID : 103993

Date : 10/10/2017 4:25:54 PM

2B 90716 1B 90716

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/13/2017	Solid	I5775-03	Corrosivity	Cool 4 deg C	PSEG05	H61	SU-04-101017	10/10/2017	9045C
10/13/2017	Solid	I5780-04	Corrosivity	Cool 4 deg C	PSEG01	I43	PUMPING-PLANT-1-5-COMP	10/10/2017	9045C
10/13/2017	Solid	I5780-08	Corrosivity	Cool 4 deg C	PSEG01	I43	DUCT-BANK-1-8-COMP	10/10/2017	9045C

Date/Time 10/10/17 4:25 PM
 Received by: JD
 Relinquished by: CD

Date/Time 10/18/17 6:35 PM
 Received by: CD
 Relinquished by: JD