


7470A

LB90735
INSTRUMENT ID: CX4

Linear

 μ Abs.:

35707



A= 0.0000e+000

B= 2.8113e-004

C= 4.3446e-003

Rho= 0.9999651

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	-15	0.000	-15				
0.2	0.200	0.205	0.005	712	0.0 %	712				
2.5	2.500	2.522	0.022	8957	0.0 %	8957				
5.0	5.000	4.987	-0.013	17722	0.0 %	17722				
7.5	7.500	7.444	-0.056	26462	0.0 %	26462				
10.0	10.000	10.043	0.043	35707	0.0 %	35707				

LB90735 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Oct 2017 11:38:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-15	-	0.0000	7470A	PPB	11 Oct 2017 11:51:35	S	Std
0.2 - 1	SO.2	712	-	0.2000	7470A	PPB	11 Oct 2017 11:53:44	S	Std
2.5 - 1	SE.5	8957	-	2.5000	7470A	PPB	11 Oct 2017 11:55:52	S	Std
5.0 - 1	SE.5	17722	-	5.0000	7470A	PPB	11 Oct 2017 11:58:00	S	Std
7.5 - 1	SE.5	26462	-	7.5000	7470A	PPB	11 Oct 2017 12:00:08	S	Std
10.0 - 1	SE.5	35707	-	10.0000	7470A	PPB	11 Oct 2017 12:02:16	S	Std
ICV02 - 1	ICV02	13234	3.7248	-7470A	PPB	11 Oct 2017 12:04:48	U	SMPL	
ICB02 - 1	ICB02	-196	-0.0508	-7470A	PPB	11 Oct 2017 12:06:54	U	SMPL	
CCV75 - 1	CCV75	17718	4.9854	-7470A	PPB	11 Oct 2017 12:09:01	U	SMPL	
CCB75 - 1	CCB75	-163	-0.0415	-7470A	PPB	11 Oct 2017 12:11:08	U	SMPL	
CRA - 1	CRA	685	0.1969	-7470A	PPB	11 Oct 2017 12:13:14	U	SMPL	
HighStd - 1	HighStd	36429	10.2457	-7470A	PPB	11 Oct 2017 12:15:21	U	SMPL	
ChkStd - 1	ChkStd	24931	7.0132	-7470A	PPB	11 Oct 2017 12:17:28	U	SMPL	
PB103146BL - 1	PBW	-276	-0.0732	-7470A	PPB	11 Oct 2017 12:19:37	U	SMPL	
PB103146BS - 1	LCSW	13662	3.8452	-7470A	PPB	11 Oct 2017 12:21:46	U	SMPL	
PB103146TB - 1	PB103116TB	-173	-0.0443	-7470A	PPB	11 Oct 2017 12:23:53	U	SMPL	
I5763-04 - 1	220-2(2TRH)-COMP	-46	-0.0086	-7470A	PPB	11 Oct 2017 12:26:01	U	SMPL	
I5765-03 - 1	BU-03-100917	-52	-0.0103	-7470A	PPB	11 Oct 2017 12:28:09	U	SMPL	
I5766-02 - 1	RS-FRACTANK	-28	-0.0035	-7470A	PPB	11 Oct 2017 12:30:18	U	SMPL	
I5766-07 - 1	RS-CONCRETE	-16	-0.0002	-7470A	PPB	11 Oct 2017 12:32:25	U	SMPL	
CCV76 - 1	CCV76	17882	5.0315	-7470A	PPB	11 Oct 2017 12:34:32	U	SMPL	
CCB76 - 1	CCB76	-201	-0.0522	-7470A	PPB	11 Oct 2017 12:36:39	U	SMPL	
I5766-07DUP - 1	RS-CONCRETEDUP	-71	-0.0156	-7470A	PPB	11 Oct 2017 12:38:46	U	SMPL	
I5766-07MS - 1	RS-CONCRETEMS	13565	3.8179	-7470A	PPB	11 Oct 2017 12:40:53	U	SMPL	
I5766-07MSD - 1	RS-CONCRETEMSD	12552	3.5331	-7470A	PPB	11 Oct 2017 12:43:00	U	SMPL	
PB103147BL - 1	PBW	-189	-0.0488	-7470A	PPB	11 Oct 2017 12:45:06	U	SMPL	
PB103147BS - 1	LCSW	13133	3.6964	-7470A	PPB	11 Oct 2017 12:47:14	U	SMPL	
I5766-01 - 1	RS-FRACTANK	2383	0.6743	-7470A	PPB	11 Oct 2017 12:49:22	U	SMPL	
I5766-01DUP - 1	RS-FRACTANKDUP	2217	0.6276	-7470A	PPB	11 Oct 2017 12:51:30	U	SMPL	
I5766-01MS - 1	RS-FRACTANKMS	16130	4.5390	-7470A	PPB	11 Oct 2017 12:53:38	U	SMPL	
I5766-01MSD - 1	RS-FRACTANKMSD	16178	4.5525	-7470A	PPB	11 Oct 2017 12:55:47	U	SMPL	
CCV77 - 1	CCV77	17861	5.0256	-7470A	PPB	11 Oct 2017 12:57:54	U	SMPL	
CCB77 - 1	CCB77	-219	-0.0572	-7470A	PPB	11 Oct 2017 13:00:01	U	SMPL	
I5766-07LX5 - 1		-66	-0.0142	-7470A	PPB	11 Oct 2017 13:02:08	U	SMPL	
I5766-07A - 1		14072	3.9604	-7470A	PPB	11 Oct 2017 13:04:14	U	SMPL	
I5766-01LX5 - 1		291	0.0862	-7470A	PPB	11 Oct 2017 13:06:21	U	SMPL	
I5766-01A - 1		17145	4.8243	-7470A	PPB	11 Oct 2017 13:08:28	U	SMPL	
CCV78 - 1	CCV78	17589	4.9492	-7470A	PPB	11 Oct 2017 13:10:35	U	SMPL	
CCB78 - 1	CCB78	-187	-0.0482	-7470A	PPB	11 Oct 2017 13:12:42	U	SMPL	