

Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB90736

Reagent/Standard	Lot/Loq #
MBAS CAL STD, 0.5PPM	WP58694
MBAS CCV STD, 1.0PPM	WP58691
MBAS ICV-LCS std 1.0PPM	WP58706
MBAS DAILY CHECK STD, 0.01PPM	WP58707
MBAS DAILY CHECK STD, 2.0PPM	WP58708
MBAS CAL STD 2 PPM	WP58698
MBAS CAL STD 1 PPM	WP58700
MBAS CAL STD 0.1 PPM	WP58703
mbas cal std 0.05 ppm	WP58704
MBAS CAL STD 0.01 PPM	WP58705
MBAS BLANK 0 PPM	WP58687
MBAS CAL STD 0.4 PPM	WP58701
MBAS CAL STD 0.2 PPM	WP58702
Chloroform, 4X4L	W2175
MBAS Wash Solution	WP57273

Intercept: 0.0001

Slope: 0.2171

Regression: 1

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Anal Date	Anal Time
1	CAL1	0	1	0.000	10/11/2017	11:47
2	CAL2	0.01	1	0.002	10/11/2017	11:51
3	CAL3	0.05	1	0.011	10/11/2017	11:52
4	CAL4	0.1	1	0.023	10/11/2017	11:52
5	CAL5	0.2	1	0.044	10/11/2017	11:53
6	CAL6	0.4	1	0.088	10/11/2017	11:53
7	CAL7	0.5	1	0.107	10/11/2017	11:54
8	CAL8	1	1	0.216	10/11/2017	11:54
9	CAL9	2	1	0.435	10/11/2017	11:55



Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB90736

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Result (mg/L)	AnalDate	Anal Time
1	ICV	1	1	0.212	0.976	10/11/2017	11:55
2	ICB		1	0.000	0.000	10/11/2017	11:56
3	CHK1	0.01	1	0.002	0.009	10/11/2017	11:56
4	CHK2	2	1	0.432	1.989	10/11/2017	11:57
5	CCV1	1	1	0.219	1.008	10/11/2017	11:57
6	CCB1		1	0.000	0.000	10/11/2017	11:58
7	LB90736BL		1	0.000	0.000	10/11/2017	11:58
8	LB90736BS	1	1	0.215	0.990	10/11/2017	11:59
9	I5274-29		1	0.211	0.971	10/11/2017	11:59
10	CCV2	1	1	0.217	0.999	10/11/2017	12:00
11	CCB2		1	0.000	0.000	10/11/2017	12:01