


7471A

LB90741
INSTRUMENT ID: CV4

Linear

 μ Abs.:

44186



A= 0.0000e+000

B= 2.2877e-004

C= 1.8905e-002

Rho= 0.9996573

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	3	0.000	3				
0.2	0.200	0.224	0.024	898	0.0 %	898				
2.5	2.500	2.581	0.081	11198	0.0 %	11198				
5.0	5.000	4.859	-0.141	21155	0.0 %	21155				
7.5	7.500	7.389	-0.111	32217	0.0 %	32217	0			
10.0	10.000	10.127	0.127	44186	0.0 %	44186				

LB90741 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 11 Oct 2017 15:02:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	7471A	PPB	11 Oct 2017 15:06:25	S	Std
0.2 - 1	50.2	898	-	0.2000	7471A	PPB	11 Oct 2017 15:08:42	S	Std
2.5 - 1	52.5	11198	-	2.5000	7471A	PPB	11 Oct 2017 15:11:00	S	Std
5.0 - 1	55	21155	-	5.0000	7471A	PPB	11 Oct 2017 15:13:17	S	Std
7.5 - 1	57.5	32217	-	7.5000	7471A	PPB	11 Oct 2017 15:15:35	S	Std
10.0 - 1	50	44186	-	10.0000	7471A	PPB	11 Oct 2017 15:20:26	S	Std
ICV03 - 1	ICV03	17062	3.9222	-7471A	PPB	PPB	11 Oct 2017 15:23:38	U	SMPL
ICB03 - 1	ICB03	-205	-0.0280	-7471A	PPB	PPB	11 Oct 2017 15:25:54	U	SMPL
CCV79 - 1	CCV79	21426	4.9206	-7471A	PPB	PPB	11 Oct 2017 15:28:10	U	SMPL
CCB79 - 1	CCB79	-158	-0.0172	-7471A	PPB	PPB	11 Oct 2017 15:30:26	U	SMPL
CRA - 1	CRA	854	0.2143	-7471A	PPB	PPB	11 Oct 2017 15:32:42	U	SMPL
HighStd - 1	HighStd	44770	10.2611	-7471A	PPB	PPB	11 Oct 2017 15:34:58	U	SMPL
ChkStd - 1	ChkStd	30825	7.0708	-7471A	PPB	PPB	11 Oct 2017 15:37:14	U	SMPL
PB103151BL - 1	PBS	-315	-0.0532	-7471A	PPB	PPB	11 Oct 2017 15:39:33	U	SMPL
PB103151BS - 1	LCSS	14794	3.4034	-7471A	PPB	PPB	11 Oct 2017 15:41:51	U	SMPL
I5765-01 - 1	BU-03-100917	598	0.1557	-7471A	PPB	PPB	11 Oct 2017 15:44:08	U	SMPL
I5775-01 - 1	SU-04-101017	592	0.1543	-7471A	PPB	PPB	11 Oct 2017 15:46:25	U	SMPL
I5779-01 - 1	FK-SF-03-003-A	1877	0.4483	-7471A	PPB	PPB	11 Oct 2017 15:48:43	U	SMPL
I5779-01DUP - 1	FK-SF-03-003-ADUP	2185	0.5188	-7471A	PPB	PPB	11 Oct 2017 15:51:00	U	SMPL
I5779-01MS - 1	FK-SF-03-003-AMS	18623	4.2793	-7471A	PPB	PPB	11 Oct 2017 15:53:16	U	SMPL
CCV80 - 1	CCV80	21059	4.8366	-7471A	PPB	PPB	11 Oct 2017 15:55:32	U	SMPL
CCB80 - 1	CCB80	-231	-0.0339	-7471A	PPB	PPB	11 Oct 2017 15:57:48	U	SMPL
I5779-01MSD - 1	FK-SF-03-003-AMSD	18287	4.2025	-7471A	PPB	PPB	11 Oct 2017 16:00:04	U	SMPL
I5779-03 - 1	FK-SF-03-003-B	1694	0.4064	-7471A	PPB	PPB	11 Oct 2017 16:02:20	U	SMPL
I5780-01 - 1	PUMPING-PLANT-1-5-COMP	28211	6.4728	-7471A	PPB	PPB	11 Oct 2017 16:04:35	U	SMPL
I5780-05 - 1	DUCT-BANK-1-8-COMP	85851	19.6593	-7471A	PPB	PPB	11 Oct 2017 16:06:52	U	SMPL
I5779-01LX5 - 1		371	0.1038	-7471A	PPB	PPB	11 Oct 2017 16:13:26	U	SMPL
I5779-01A - 1		18945	4.3530	-7471A	PPB	PPB	11 Oct 2017 16:15:43	U	SMPL
I5780-05DLX5 - 1		16952	3.8971	-7471A	PPB	PPB	11 Oct 2017 16:18:00	U	SMPL
CCV81 - 1	CCV81	20955	4.8128	-7471A	PPB	PPB	11 Oct 2017 16:20:18	U	SMPL
CCB81 - 1	CCB81	-236	-0.0351	-7471A	PPB	PPB	11 Oct 2017 16:22:35	U	SMPL