

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: DHVANI

Date: 10/12/2017

DATE & TIME IN: 10/12/2017 10:10

OVEN TEMP: 103 °C

Run Number: LB90757

DATE & TIME OUT: 10/12/2017 14:05

OVEN TEMP: 103 °C

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB90757BL	LB90757BL	1.4328	1.4328	100	1.4328	1.4328	1.4328	0	0
2	LB90757BS	LB90757BS	1.4307	1.4307	100	1.4854	1.4854	1.4854	0.0547	547
3	I5822-11	CLEARWELL	1.4208	1.4208	400	1.4234	1.4234	1.4234	0.0026	6.5
4	I5822-11DUP	CLEARWELLDUP	1.4366	1.4366	400	1.4391	1.4391	1.4391	0.0025	6.3
5	I5823-01	COMP	1.4232	1.4232	100	1.4423	1.4423	1.4423	0.0191	191
6	I5824-01	TWA-EFFE-GW	1.4104	1.4104	400	1.4194	1.4194	1.4194	0.009	22.5
7	I5824-02	TWA-EFFT2.2-GW	1.4273	1.4273	400	1.4401	1.4401	1.4401	0.0128	32

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

LB90757

WorkList Name : TSS101217

WorkList ID : 104050

Date : 10/12/2017 8:32:16 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/18/2017	Water	I5822-11 ✓	TSS	Cool 4 deg C	CARP01	H11	CLEARWELL	10/11/2017	SM2540 D
10/18/2017	Water	I5823-01 ✓	TSS	Cool 4 deg C	ARAM01	H11	COMP	10/11/2017	SM2540 D
10/12/2017	Water	I5824-01 ✓	TSS	Cool 4 deg C	LIRO01	J23	TWA-EFFE-GW	10/11/2017	SM2540 D
10/12/2017	Water	I5824-02 ✓	TSS	Cool 4 deg C	LIRO01	J23	TWA-EFFT2.2-GW	10/11/2017	SM2540 D

Date/Time 10/12/17 9:45
 Received by: DP
 Relinquished by: CP

Date/Time 10/12/17 10:20
 Received by: CP
 Relinquished by: DP