

LB 90758


7470A

INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

34754



A= 0.0000e+000

B= 2.8371e-004

C= 2.1529e-002

Rho= 0.9997478

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.007 | 0.007  | -50   | 0.000      | -50   |       |       |       |       |
| 0.2    | 0.200  | 0.186 | -0.014 | 581   | 0.0 %      | 581   |       |       |       |       |
| 2.5    | 2.500  | 2.438 | -0.062 | 8518  | 0.0 %      | 8518  |       |       |       |       |
| 5.0    | 5.000  | 5.040 | 0.040  | 17687 | 0.0 %      | 17687 |       |       |       |       |
| 7.5    | 7.500  | 7.647 | 0.147  | 26877 | 0.0 %      | 26877 |       |       |       |       |
| 10.0   | 10.000 | 9.882 | -0.118 | 34754 | 0.0 %      | 34754 |       |       |       |       |

## LB90758 INSTRUMENT ID : CV1

| Method: 7470A   |                        | Operator: Admin |         | Date of Analysis: 12 Oct 2017 09:33:50 |       |                      |      |      |
|-----------------|------------------------|-----------------|---------|----------------------------------------|-------|----------------------|------|------|
| Sample ID       | Extended ID            | μ Abs.          | Conc.   | Std ConcMethod                         | Units | Date                 | Type | Type |
| 0.0 - 1         | 50                     | -50             | -       | 0.00007470A                            | PPB   | 12 Oct 2017 09:39:46 | S    | Std  |
| 0.2 - 1         | 50.2                   | 581             | -       | 0.20007470A                            | PPB   | 12 Oct 2017 09:41:54 | S    | Std  |
| 2.5 - 1         | 52.5                   | 8518            | -       | 2.50007470A                            | PPB   | 12 Oct 2017 09:44:02 | S    | Std  |
| 5.0 - 1         | 55                     | 17687           | -       | 5.00007470A                            | PPB   | 12 Oct 2017 09:46:11 | S    | Std  |
| 7.5 - 1         | 57.5                   | 26877           | -       | 7.50007470A                            | PPB   | 12 Oct 2017 09:48:18 | S    | Std  |
| 10.0 - 1        | 510                    | 34754           | -       | 10.00007470A                           | PPB   | 12 Oct 2017 09:50:26 | S    | Std  |
| ICV04 - 1       | ICV04                  | 13524           | 3.8585  | -7470A                                 | PPB   | 12 Oct 2017 09:53:03 | U    | SMPL |
| ICB04 - 1       | ICB04                  | -207            | -0.0372 | -7470A                                 | PPB   | 12 Oct 2017 09:55:09 | U    | SMPL |
| CCV82 - 1       | CCV82                  | 17414           | 4.9621  | -7470A                                 | PPB   | 12 Oct 2017 09:57:15 | U    | SMPL |
| CCB82 - 1       | CCB82                  | -256            | -0.0511 | -7470A                                 | PPB   | 12 Oct 2017 09:59:22 | U    | SMPL |
| CRA - 1         | CRA                    | 554             | 0.1787  | -7470A                                 | PPB   | 12 Oct 2017 10:01:29 | U    | SMPL |
| HighStd - 1     | HighStd                | 34826           | 9.9021  | -7470A                                 | PPB   | 12 Oct 2017 10:03:36 | U    | SMPL |
| ChkStd - 1      | ChkStd                 | 23479           | 6.6828  | -7470A                                 | PPB   | 12 Oct 2017 10:05:43 | U    | SMPL |
| PB103182BL - 1  | PBW                    | -254            | -0.0505 | -7470A                                 | PPB   | 12 Oct 2017 10:07:52 | U    | SMPL |
| PB103182BS - 1  | LCSW                   | 13394           | 3.8216  | -7470A                                 | PPB   | 12 Oct 2017 10:10:00 | U    | SMPL |
| PB103182TB - 1  | PB103148TB             | -107            | -0.0088 | -7470A                                 | PPB   | 12 Oct 2017 10:12:08 | U    | SMPL |
| I5775-03 - 1    | SU-04-101017           | -32             | 0.0124  | -7470A                                 | PPB   | 12 Oct 2017 10:14:16 | U    | SMPL |
| I5780-04 - 1    | PUMPING-PLANT-1-5-COMP | 436             | 0.1452  | -7470A                                 | PPB   | 12 Oct 2017 10:16:24 | U    | SMPL |
| I5780-08 - 1    | DUCT-BANK-1-8-COMP     | 114             | 0.0539  | -7470A                                 | PPB   | 12 Oct 2017 10:18:33 | U    | SMPL |
| I5780-08DUP - 1 | DUCT-BANK-1-8-COMPDUP  | 87              | 0.0462  | -7470A                                 | PPB   | 12 Oct 2017 10:20:39 | U    | SMPL |
| CCV83 - 1       | CCV83                  | 17323           | 4.9363  | -7470A                                 | PPB   | 12 Oct 2017 10:22:46 | U    | SMPL |
| CCB83 - 1       | CCB83                  | -162            | -0.0244 | -7470A                                 | PPB   | 12 Oct 2017 10:24:53 | U    | SMPL |
| I5780-08MS - 1  | DUCT-BANK-1-8-COMPMS   | 13080           | 3.7325  | -7470A                                 | PPB   | 12 Oct 2017 10:27:00 | U    | SMPL |
| I5780-08MSD - 1 | DUCT-BANK-1-8-COMPMSD  | 12784           | 3.6485  | -7470A                                 | PPB   | 12 Oct 2017 10:29:07 | U    | SMPL |
| I5780-08LX5 - 1 |                        | -142            | -0.0188 | -7470A                                 | PPB   | 12 Oct 2017 10:31:14 | U    | SMPL |
| I5780-08A - 1   |                        | 14132           | 4.0310  | -7470A                                 | PPB   | 12 Oct 2017 10:33:21 | U    | SMPL |
| CCV84 - 1       | CCV84                  | 17688           | 5.0398  | -7470A                                 | PPB   | 12 Oct 2017 10:35:28 | U    | SMPL |
| CCB84 - 1       | CCB84                  | -198            | -0.0346 | -7470A                                 | PPB   | 12 Oct 2017 10:37:36 | U    | SMPL |