


245.1

LB90765
INSTRUMENT ID: 042

Linear

 μ Abs.:

31888



A= 0.0000e+000

B= 3.1315e-004

C= -1.6985e-002

Rho= 0.9999114

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.025	-0.025	-25	0.000	-25				
0.2	0.200	0.198	-0.002	688	0.0 %	688				
2.5	2.500	2.562	0.062	8234	0.0 %	8234				
5.0	5.000	4.931	-0.069	15799	0.0 %	15799				
7.5	7.500	7.565	0.065	24213	0.0 %	24213				
10.0	10.000	9.969	-0.031	31888	0.0 %	31888				

LB90765 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Oct 2017 10:52:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		-25	-	0.0000	245.1	PPB	12 Oct 2017 11:06:09	S	Std
0.2 - 1		688	-	0.2000	245.1	PPB	12 Oct 2017 11:08:17	S	Std
2.5 - 1		8234	-	2.5000	245.1	PPB	12 Oct 2017 11:10:26	S	Std
5.0 - 1		15799	-	5.0000	245.1	PPB	12 Oct 2017 11:12:34	S	Std
7.5 - 1		24213	-	7.5000	245.1	PPB	12 Oct 2017 11:14:42	S	Std
10.0 - 1		31888	-	10.0000	245.1	PPB	12 Oct 2017 11:16:50	S	Std
ICV05 - 1	ICV05	12705	3.9616	-245.1		PPB	12 Oct 2017 11:19:28	U	SMPL
ICB05 - 1	ICB05	-145	-0.0624	-245.1		PPB	12 Oct 2017 11:21:35	U	SMPL
CCV85 - 1	CCV85	15981	4.9875	-245.1		PPB	12 Oct 2017 11:23:41	U	SMPL
CCB85 - 1	CCB85	-175	-0.0718	-245.1		PPB	12 Oct 2017 11:25:48	U	SMPL
CRA - 1	CRA	661	0.1900	-245.1		PPB	12 Oct 2017 11:27:56	U	SMPL
HighStd - 1	HighStd	32036	10.0152	-245.1		PPB	12 Oct 2017 11:30:02	U	SMPL
ChkStd - 1	ChkStd	22113	6.9078	-245.1		PPB	12 Oct 2017 11:32:09	U	SMPL
PB103170BL - 1	PBW	-257	-0.0975	-245.1		PPB	12 Oct 2017 11:34:18	U	SMPL
PB103170BS - 1	LCSW	13036	4.0653	-245.1		PPB	12 Oct 2017 11:36:27	U	SMPL
I5777-01 - 1	100917-TIDO-F-D-B	-254	-0.0965	-245.1		PPB	12 Oct 2017 11:38:34	U	SMPL
I5777-02 - 1	100917-TIDO-F-D-S	-130	-0.0577	-245.1		PPB	12 Oct 2017 11:40:42	U	SMPL
I5777-03 - 1	100917-TIDO-F-U-B	-106	-0.0502	-245.1		PPB	12 Oct 2017 11:42:51	U	SMPL
I5777-04 - 1	100917-TIDO-F-U-S	-87	-0.0442	-245.1		PPB	12 Oct 2017 11:45:00	U	SMPL
I5777-04DUP - 1	100917-TIDO-F-U-SDUP	-64	-0.0370	-245.1		PPB	12 Oct 2017 11:47:06	U	SMPL
CCV86 - 1	CCV86	15835	4.9418	-245.1		PPB	12 Oct 2017 11:49:12	U	SMPL
CCB86 - 1	CCB86	-204	-0.0809	-245.1		PPB	12 Oct 2017 11:51:19	U	SMPL
I5777-04MS - 1	100917-TIDO-F-U-SMS	14934	4.6597	-245.1		PPB	12 Oct 2017 11:53:27	U	SMPL
I5777-04MSD - 1	100917-TIDO-F-U-SMSD	14838	4.6296	-245.1		PPB	12 Oct 2017 11:55:34	U	SMPL
I5777-04LX5 - 1		-132	-0.0583	-245.1		PPB	12 Oct 2017 11:57:41	U	SMPL
I5777-04A - 1		16064	5.0135	-245.1		PPB	12 Oct 2017 11:59:48	U	SMPL
CCV87 - 1	CCV87	15738	4.9114	-245.1		PPB	12 Oct 2017 12:01:55	U	SMPL
CCB87 - 1	CCB87	-131	-0.0580	-245.1		PPB	12 Oct 2017 12:04:03	U	SMPL