

LB 90770


7471A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44267



A= 0.0000e+000

B= 2.2642e-004

C= 1.7167e-002

Rho= 0.9999721

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	-81	0.000	-81				
0.2	0.200	0.227	0.027	927	0.0 %	927				
2.5	2.500	2.500	0.000	10967	0.0 %	10967				
5.0	5.000	4.966	-0.034	21856	0.0 %	21856				
7.5	7.500	7.468	-0.032	32905	0.0 %	32905	0			
10.0	10.000	10.040	0.040	44267	0.0 %	44267				

LB90770 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 12 Oct 2017 12:16:28				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-81	-	0.00007471A	PPB	12 Oct 2017 12:19:41	S	Std
0.2 - 1	50.2	927	-	0.20007471A	PPB	12 Oct 2017 12:21:59	S	Std
2.5 - 1	2.5	10967	-	2.50007471A	PPB	12 Oct 2017 12:24:17	S	Std
5.0 - 1	55	21856	-	5.00007471A	PPB	12 Oct 2017 12:26:35	S	Std
7.5 - 1	57.5	32905	-	7.50007471A	PPB	12 Oct 2017 12:28:52	S	Std
10.0 - 1	510	44267	-	10.00007471A	PPB	12 Oct 2017 12:31:09	S	Std
ICV06 - 1	ICV06	16253	3.6972	-7471A	PPB	12 Oct 2017 12:33:52	U	SMPL
ICB06 - 1	ICB06	-138	-0.0141	-7471A	PPB	12 Oct 2017 12:36:08	U	SMPL
CCV88 - 1	CCV88	21930	4.9826	-7471A	PPB	12 Oct 2017 12:38:24	U	SMPL
CCB88 - 1	CCB88	-145	-0.0157	-7471A	PPB	12 Oct 2017 12:40:40	U	SMPL
CRA - 1	CRA	1001	0.2438	-7471A	PPB	12 Oct 2017 12:42:57	U	SMPL
HighStd - 1	HighStd	44884	10.1800	-7471A	PPB	12 Oct 2017 12:45:13	U	SMPL
ChkStd - 1	ChkStd	31244	7.0915	-7471A	PPB	12 Oct 2017 12:47:29	U	SMPL
PB103184BL - 1	PBS	-287	-0.0478	-7471A	PPB	12 Oct 2017 12:49:47	U	SMPL
PB103184BS - 1	LCSS	16128	3.6689	-7471A	PPB	12 Oct 2017 12:54:51	U	SMPL
I5776-01 - 1	ENV-114-2(0-5)	26142	5.9363	-7471A	PPB	12 Oct 2017 12:57:09	U	SMPL
I5776-02 - 1	ENV-114-2(5-10)	16081	3.6583	-7471A	PPB	12 Oct 2017 12:59:27	U	SMPL
I5776-03 - 1	ENV-114-2(10-15)	90	0.0375	-7471A	PPB	12 Oct 2017 13:01:45	U	SMPL
I5776-04 - 1	ENV-107-1(0-5)	963	0.2352	-7471A	PPB	12 Oct 2017 13:04:03	U	SMPL
I5776-05 - 1	ENV-104-2(0-5)	1727	0.4082	-7471A	PPB	12 Oct 2017 13:06:20	U	SMPL
CCV89 - 1	CCV89	21934	4.9835	-7471A	PPB	12 Oct 2017 13:08:36	U	SMPL
CCB89 - 1	CCB89	-139	-0.0143	-7471A	PPB	12 Oct 2017 13:10:52	U	SMPL
I5776-06 - 1	ENV-104-2(5-10)	1596	0.3785	-7471A	PPB	12 Oct 2017 13:13:08	U	SMPL
I5776-07 - 1	EPE-P6(50-60)	758	0.1888	-7471A	PPB	12 Oct 2017 13:15:25	U	SMPL
I5776-08 - 1	EPE-P6(60-70)	850	0.2096	-7471A	PPB	12 Oct 2017 13:17:41	U	SMPL
I5776-09 - 1	EPE-P6(70-80)	533	0.1379	-7471A	PPB	12 Oct 2017 13:19:57	U	SMPL
I5776-10 - 1	EPE-P6(80-90)	661	0.1668	-7471A	PPB	12 Oct 2017 13:22:14	U	SMPL
I5782-01 - 1	S1(1-3)	29038	6.5921	-7471A	PPB	12 Oct 2017 13:24:31	U	SMPL
I5782-01DUP - 1	S1(1-3)DUP	29832	6.7718	-7471A	PPB	12 Oct 2017 13:26:48	U	SMPL
I5782-01MS - 1	S1(1-3)MS	49665	11.2625	-7471A	PPB	12 Oct 2017 13:29:08	U	SMPL
I5782-01MSD - 1	S1(1-3)MSD	47524	10.7777	-7471A	PPB	12 Oct 2017 13:31:26	U	SMPL
I5782-02 - 1	S1(10-12)	10696	2.4390	-7471A	PPB	12 Oct 2017 13:33:44	U	SMPL
CCV90 - 1	CCV90	21968	4.9912	-7471A	PPB	12 Oct 2017 13:36:02	U	SMPL
CCB90 - 1	CCB90	-162	-0.0195	-7471A	PPB	12 Oct 2017 13:38:19	U	SMPL
I5782-03 - 1	S2(1-3)	44463	10.0846	-7471A	PPB	12 Oct 2017 13:40:35	U	SMPL
I5782-04 - 1	S2(10-12)	5287	1.2143	-7471A	PPB	12 Oct 2017 13:42:51	U	SMPL
I5782-05 - 1	S3(1-3)	169474	38.3901	-7471A	PPB	12 Oct 2017 13:45:10	U	SMPL
I5782-06 - 1	S3(10-12)	-3	0.0165	-7471A	PPB	12 Oct 2017 13:52:34	U	SMPL
I5782-07 - 1	S4(1-3)	6973	1.5960	-7471A	PPB	12 Oct 2017 13:54:50	U	SMPL
I5782-08 - 1	S4(10-12)	44	0.0271	-7471A	PPB	12 Oct 2017 13:57:07	U	SMPL
I5782-09 - 1	S5(1-3)	8148	1.8621	-7471A	PPB	12 Oct 2017 13:59:24	U	SMPL
I5782-10 - 1	S5(10-12)	1907	0.4490	-7471A	PPB	12 Oct 2017 14:01:42	U	SMPL
PB103189BL - 1	PBS	-24	0.0117	-7471A	PPB	12 Oct 2017 14:04:00	U	SMPL
PB103189BS - 1	LCSS	16015	3.6433	-7471A	PPB	12 Oct 2017 14:06:16	U	SMPL
CCV91 - 1	CCV91	21993	4.9969	-7471A	PPB	12 Oct 2017 14:08:33	U	SMPL
CCB91 - 1	CCB91	-170	-0.0213	-7471A	PPB	12 Oct 2017 14:10:49	U	SMPL
I5794-01 - 1	ENV-106-5(0-5)	7979	1.8238	-7471A	PPB	12 Oct 2017 14:18:11	U	SMPL
I5794-01DUP - 1	ENV-106-5(0-5)DUP	7494	1.7140	-7471A	PPB	12 Oct 2017 14:20:27	U	SMPL
I5794-01MS - 1	ENV-106-5(0-5)MS	25252	5.7348	-7471A	PPB	12 Oct 2017 14:25:42	U	SMPL
I5794-01MSD - 1	ENV-106-5(0-5)MSD	25075	5.6947	-7471A	PPB	12 Oct 2017 14:27:58	U	SMPL
I5794-02 - 1	ENV-106-5(0-10)	1024	0.2490	-7471A	PPB	12 Oct 2017 14:30:16	U	SMPL
I5794-03 - 1	ENV-106-5(10-15)	87	0.0369	-7471A	PPB	12 Oct 2017 14:32:34	U	SMPL
I5794-04 - 1	ENV-106-5(15-20)	149	0.0509	-7471A	PPB	12 Oct 2017 14:34:52	U	SMPL
I5794-05 - 1	ENV-104-1(0-5)	578	0.1480	-7471A	PPB	12 Oct 2017 14:37:09	U	SMPL
I5794-06 - 1	ENV-104-1(5-10)	375	0.1021	-7471A	PPB	12 Oct 2017 14:39:27	U	SMPL
I5794-07 - 1	ENV-105-4(0-5)	3426	0.7929	-7471A	PPB	12 Oct 2017 14:41:43	U	SMPL
CCV92 - 1	CCV92	22109	5.0232	-7471A	PPB	12 Oct 2017 14:44:00	U	SMPL
CCB92 - 1	CCB92	-236	-0.0363	-7471A	PPB	12 Oct 2017 14:46:16	U	SMPL
I5794-08 - 1	ENV-105-4(5-10)	169	0.0554	-7471A	PPB	12 Oct 2017 14:48:33	U	SMPL
I5794-09 - 1	ENV-105-4(10-15)	77	0.0346	-7471A	PPB	12 Oct 2017 14:50:50	U	SMPL
I5794-10 - 1	EPE-117-3(10-15)	297	0.0844	-7471A	PPB	12 Oct 2017 14:53:06	U	SMPL
I5813-01 - 1	OK-03-101117	5709	1.3098	-7471A	PPB	12 Oct 2017 14:55:22	U	SMPL
I5816-01 - 1	OR-01-101117	822	0.2033	-7471A	PPB	12 Oct 2017 14:57:39	U	SMPL
I5820-01 - 1	SOILPILE	16859	3.8344	-7471A	PPB	12 Oct 2017 14:59:55	U	SMPL
I5821-01 - 1	WPSB-5A-2-BGS	10484	2.3910	-7471A	PPB	12 Oct 2017 15:02:13	U	SMPL
I5821-02 - 1	WPSB-5B-23.5-24-BGS	70	0.0330	-7471A	PPB	12 Oct 2017 15:04:30	U	SMPL
I5821-03 - 1	WPSB-6A-2-BGS	20817	4.7306	-7471A	PPB	12 Oct 2017 15:06:48	U	SMPL
I5821-04 - 1	WPSB-6B-18-18.5-BGS	58	0.0303	-7471A	PPB	12 Oct 2017 15:09:05	U	SMPL
CCV93 - 1	CCV93	22219	5.0481	-7471A	PPB	12 Oct 2017 15:11:22	U	SMPL
CCB93 - 1	CCB93	-243	-0.0379	-7471A	PPB	12 Oct 2017 15:13:39	U	SMPL
I5821-05 - 1	WPSB-7A-2-BGS	56743	12.8651	-7471A	PPB	12 Oct 2017 15:15:56	U	SMPL
I5821-06 - 1	WPSB-7B-16.5-17-BGS	98	0.0394	-7471A	PPB	12 Oct 2017 15:18:13	U	SMPL
I5782-01LX5 - 1		5725	1.3134	-7471A	PPB	12 Oct 2017 15:20:32	U	SMPL
I5782-01A - 1		46033	10.4401	-7471A	PPB	12 Oct 2017 15:22:48	U	SMPL
I5794-01LX5 - 1		1485	0.3534	-7471A	PPB	12 Oct 2017 15:32:41	U	SMPL
I5794-01A - 1		24062	5.4654	-7471A	PPB	12 Oct 2017 15:34:58	U	SMPL
CCV94 - 1	CCV94	22073	5.0150	-7471A	PPB	12 Oct 2017 15:37:16	U	SMPL
CCB94 - 1	CCB94	-239	-0.0369	-7471A	PPB	12 Oct 2017 15:39:35	U	SMPL
I5782-01DLX2 - 1		14710	3.3479	-7471A	PPB	12 Oct 2017 15:41:54	U	SMPL
I5782-01DUPDLX2 - 1		14835	3.3762	-7471A	PPB	12 Oct 2017 15:44:10	U	SMPL
I5782-01MSDLX2 - 1		24779	5.6277	-7471A	PPB	12 Oct 2017 15:46:27	U	SMPL

LB90770 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 12 Oct 2017 12:16:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
I5782-01MSDDLX2 - 1		23635	5.3687		-7471A	PPB	12 Oct 2017 15:48:44	U	SMPL
I5782-01LDLX10 - 1		2829	0.6577		-7471A	PPB	12 Oct 2017 15:51:02	U	SMPL
I5782-01ADLX2 - 1		24479	5.5598		-7471A	PPB	12 Oct 2017 15:57:20	U	SMPL
I5782-03DLX2 - 1		22039	5.0073		-7471A	PPB	12 Oct 2017 15:59:37	U	SMPL
I5782-05DLX10 - 1		17652	4.0140		-7471A	PPB	12 Oct 2017 16:01:55	U	SMPL
5821-05DLX2 - 1		27718	6.2932		-7471A	PPB	12 Oct 2017 16:04:12	U	SMPL
CCV95 - 1	CCV95	21741	4.9398		-7471A	PPB	12 Oct 2017 16:06:30	U	SMPL
CCB95 - 1	CCB95	-224	-0.0336		-7471A	PPB	12 Oct 2017 16:08:48	U	SMPL