



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90772

Temp.: 21.3 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/12/2017	14:00
2	CAL2	1	Water			7.00	10/12/2017	14:01
3	CAL3	1	Water			10.02	10/12/2017	14:02
4	ICV	1	Water			7.00	10/12/2017	14:05
5	CCV1	1	Water			2.02	10/12/2017	14:06
6	I5817-07	1	Water			5.80	10/12/2017	14:07
7	I5817-07DUP	1	Water			5.81	10/12/2017	14:08
8	I5817-08	1	Water			6.02	10/12/2017	14:10
9	I5817-09	1	Water			7.31	10/12/2017	14:11
10	I5817-10	1	Water			6.75	10/12/2017	14:13
11	CCV2	1	Water			12.01	10/12/2017	14:15

2390772

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5717 WorkList ID : 104095 Date : 10/12/2017 1:46:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2017	Water	I5817-07	Corrosivity	Cool 4 deg C	DVIR01	J61	HW-1	10/11/2017	9040C
10/11/2017	Water	I5817-08	Corrosivity	Cool 4 deg C	DVIR01	J61	HW-32	10/11/2017	9040C
10/11/2017	Water	I5817-09	Corrosivity	Cool 4 deg C	DVIR01	J61	TANK-3	10/11/2017	9040C
10/11/2017	Water	I5817-10	Corrosivity	Cool 4 deg C	DVIR01	J61	HW-3	10/11/2017	9040C

Date/Time 10/12/17 11:46 AM
Received by: [Signature]
Relinquished by:

Date/Time 10/12/17 3:10 PM
Received by: [Signature]
Relinquished by: JP