



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB90773

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	10/12/2017	09:10
2	I5766-02		1	>100.0	>212.0	10/12/2017	09:42
3	I5766-02DUP		1	>100.0	>212.0	10/12/2017	10:15
4	I5817-07		1	>100.0	>212.0	10/12/2017	10:42
5	I5817-08		1	>100.0	>212.0	10/12/2017	11:12
6	I5817-09		1	>100.0	>212.0	10/12/2017	11:35
7	I5817-10		1	>100.0	>212.0	10/12/2017	12:02

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP101217

WorkList ID : 104096

Date : 10/12/2017 0830AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/12/2017	Water	I5766-02	Flash Point	Cool 4 deg C	PSEG01	H53	RS-FRACTANK	10/09/2017	1010A
10/24/2017	Water	I5817-07	Flash Point	Cool 4 deg C	DVIR01	J61	HW-1	10/11/2017	1010A
10/24/2017	Water	I5817-08	Flash Point	Cool 4 deg C	DVIR01	J61	HW-32	10/11/2017	1010A
10/24/2017	Water	I5817-09	Flash Point	Cool 4 deg C	DVIR01	J61	TANK-3	10/11/2017	1010A
10/24/2017	Water	I5817-10	Flash Point	Cool 4 deg C	DVIR01	J61	HW-3	10/11/2017	1010A

Date/Time

Received by:

Relinquished by:

10/12/17 8:35AM
RS
CP

Date/Time

Received by:

Relinquished by:

10/12/17 12:10PM
CP
KS