



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90776

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/12/2017	16:25
2	CAL2	1	Water			7.01	10/12/2017	16:26
3	CAL3	1	Water			10.03	10/12/2017	16:27
4	ICV	1	Water			7.01	10/12/2017	16:30
5	CCV1	1	Water			2.01	10/12/2017	16:31
6	I5857-01	1	Solid	20.02	20	5.81	10/12/2017	16:32
7	I5857-01DUP	1	Solid	20.03	20	5.82	10/12/2017	16:34
8	CCV2	1	Water			12.01	10/12/2017	16:35

2B90776

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5857

WorkList ID : 104099

Date : 10/12/2017 3:52:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/12/2017	Solid	I5857-01	pH	Cool 4 deg C	TULL02	I13	FK-PS-01-018	10/12/2017	9045C

Date/Time 10/12/17 6:00 PM
Received by: JF
Relinquished by: CR

Date/Time 10/12/17 5:00 PM
Received by: CR
Relinquished by: JF