



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90779

Temp.: 19.6 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/12/2017	17:20
2	CAL2	1	Water			7.00	10/12/2017	17:21
3	CAL3	1	Water			10.03	10/12/2017	17:22
4	ICV	1	Water			6.99	10/12/2017	17:24
5	CCV1	1	Water			2.01	10/12/2017	17:25
6	I5875-01	1	Water			7.71	10/12/2017	17:26
7	I5875-01DUP	1	Water			7.72	10/12/2017	17:28
8	CCV2	1	Water			12.01	10/12/2017	17:30

2890779

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5875-1

WorkList ID : 104102

Date : 10/12/2017 5:04:19 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/12/2017	Water	I5875-01-17	pH	Cool 4 deg C	PSEG01	H52	TK2064L-T1-T2-T3	10/12/2017	9045C

Date/Time 10-12-17 5:05 PM
Received by: m
Relinquished by: u

Date/Time 10-12-17 6:15 PM
Received by: u
Relinquished by: JP