


7470A

LB 90796
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35049



A= 0.0000e+000

B= 2.8939e-004

C= -1.1458e-002

Rho= 0.9996548

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	74	0.000	74				
0.2	0.200	0.211	0.011	768	0.0 %	768				
2.5	2.500	2.557	0.057	8876	0.0 %	8876				
5.0	5.000	4.978	-0.022	17242	0.0 %	17242				
7.5	7.500	7.312	-0.188	25308	0.0 %	25308				
10.0	10.000	10.131	0.131	35049	0.0 %	35049				

LB90796 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Oct 2017 10:38:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	74	-	0.00007470A	PPB		13 Oct 2017 11:14:08	S	Std
0.2 - 1	50.2	768	-	0.20007470A	PPB		13 Oct 2017 11:16:16	S	Std
2.5 - 1	52.5	8876	-	2.50007470A	PPB		13 Oct 2017 11:18:25	S	Std
5.0 - 1	55	17242	-	5.00007470A	PPB		13 Oct 2017 11:20:33	S	Std
7.5 - 1	57.5	25308	-	7.50007470A	PPB		13 Oct 2017 11:22:41	S	Std
10.0 - 1	510	35049	-	10.00007470A	PPB		13 Oct 2017 11:27:16	S	Std
ICV07 - 1	ICV07	13450	3.8809	-7470A	PPB		13 Oct 2017 11:29:48	U	SMPL
ICB07 - 1	ICB07	-137	-0.0511	-7470A	PPB		13 Oct 2017 11:31:55	U	SMPL
CCV01 - 1	CCV01	17390	5.0211	-7470A	PPB		13 Oct 2017 11:34:02	U	SMPL
CCB01 - 1	CCB01	-145	-0.0534	-7470A	PPB		13 Oct 2017 11:36:08	U	SMPL
CRA - 1	CRA	674	0.1836	-7470A	PPB		13 Oct 2017 11:38:15	U	SMPL
HighStd - 1	HighStd	35246	10.1884	-7470A	PPB		13 Oct 2017 11:40:21	U	SMPL
ChkStd - 1	ChkStd	24069	6.9539	-7470A	PPB		13 Oct 2017 11:42:28	U	SMPL
PB103225BL - 1	PBW	-231	-0.0783	-7470A	PPB		13 Oct 2017 11:44:38	U	SMPL
PB103225BS - 1	LCSW	13494	3.8936	-7470A	PPB		13 Oct 2017 11:46:46	U	SMPL
PB103225TB - 1	PB103187TB	-101	-0.0407	-7470A	PPB		13 Oct 2017 11:48:53	U	SMPL
I5791-01 - 1	CLEANUP-DRUM	-13	-0.0152	-7470A	PPB		13 Oct 2017 11:51:02	U	SMPL
I5810-01 - 1	CB-1017-S-TCLP-43S	31	-0.0025	-7470A	PPB		13 Oct 2017 11:53:10	U	SMPL
I5812-01 - 1	DRUM-1	12	-0.0080	-7470A	PPB		13 Oct 2017 11:55:19	U	SMPL
I5812-02 - 1	DRUM-2	5	-0.0100	-7470A	PPB		13 Oct 2017 11:57:26	U	SMPL
CCV02 - 1	CCV02	17305	4.9965	-7470A	PPB		13 Oct 2017 11:59:33	U	SMPL
CCB02 - 1	CCB02	-126	-0.0479	-7470A	PPB		13 Oct 2017 12:01:39	U	SMPL
I5813-03 - 1	OK-03-101117	-24	-0.0184	-7470A	PPB		13 Oct 2017 12:03:46	U	SMPL
I5816-03 - 1	OR-01-101117	-24	-0.0184	-7470A	PPB		13 Oct 2017 12:05:53	U	SMPL
I5816-03DUP - 1	OR-01-101117DUP	-24	-0.0184	-7470A	PPB		13 Oct 2017 12:07:59	U	SMPL
I5816-03MS - 1	OR-01-101117MS	13115	3.7839	-7470A	PPB		13 Oct 2017 12:10:06	U	SMPL
I5816-03MSD - 1	OR-01-101117MSD	12768	3.6835	-7470A	PPB		13 Oct 2017 12:12:13	U	SMPL
I5816-03LX5 - 1		-100	-0.0404	-7470A	PPB		13 Oct 2017 12:14:21	U	SMPL
I5816-03A - 1		13964	4.0296	-7470A	PPB		13 Oct 2017 12:16:29	U	SMPL
CCV03 - 1	CCV03	16948	4.8931	-7470A	PPB		13 Oct 2017 12:18:37	U	SMPL
CCB03 - 1	CCB03	-168	-0.0601	-7470A	PPB		13 Oct 2017 12:20:45	U	SMPL