

LB 90802


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

31303



A= 0.0000e+000

B= 3.1922e-004

C= -4.4023e-002

Rho= 0.9999569

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.030	-0.030	45	0.000	45				
0.2	0.200	0.197	-0.003	756	0.0 %	756				
2.5	2.500	2.501	0.001	7972	0.0 %	7972				
5.0	5.000	5.047	0.047	15947	0.0 %	15947				
7.5	7.500	7.536	0.036	23747	0.0 %	23747				
10.0	10.000	9.949	-0.051	31303	0.0 %	31303				

LB90802 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Oct 2017 14:06:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000	245.1	PPB	13 Oct 2017 14:08:05	S	Std
0.2 - 1	50.2	756	-	0.2000	245.1	PPB	13 Oct 2017 14:10:13	S	Std
2.5 - 1	52.5	7972	-	2.5000	245.1	PPB	13 Oct 2017 14:12:22	S	Std
5.0 - 1	55	15947	-	5.0000	245.1	PPB	13 Oct 2017 14:14:30	S	Std
7.5 - 1	57.5	23747	-	7.5000	245.1	PPB	13 Oct 2017 14:16:38	S	Std
10.0 - 1	510	31303	-	10.0000	245.1	PPB	13 Oct 2017 14:23:34	S	Std
ICV08 - 1	ICV08	12100	3.8185	-245.1		PPB	13 Oct 2017 14:26:07	U	SMPL
ICB08 - 1	ICB08	-174	-0.0996	-245.1		PPB	13 Oct 2017 14:28:13	U	SMPL
CCV04 - 1	CCV04	15572	4.9269	-245.1		PPB	13 Oct 2017 14:30:20	U	SMPL
CCB04 - 1	CCB04	-143	-0.0897	-245.1		PPB	13 Oct 2017 14:32:27	U	SMPL
CRA - 1	CRA	721	0.1861	-245.1		PPB	13 Oct 2017 14:34:34	U	SMPL
HighStd - 1	HighStd	31895	10.1375	-245.1		PPB	13 Oct 2017 14:36:41	U	SMPL
ChkStd - 1	ChkStd	21883	6.9415	-245.1		PPB	13 Oct 2017 14:38:48	U	SMPL
PB103226BL - 1	PBW	-170	-0.0983	-245.1		PPB	13 Oct 2017 14:40:57	U	SMPL
PB103226BS - 1	LCSW	14089	4.4535	-245.1		PPB	13 Oct 2017 14:43:05	U	SMPL
I5686-01 - 1	EFFLUENT-DAY1METALS	57	-0.0258	-245.1		PPB	13 Oct 2017 14:45:12	U	SMPL
I5686-01DUP - 1	EFFLUENT-DAY1METALSDUP	-93	-0.0737	-245.1		PPB	13 Oct 2017 14:47:21	U	SMPL
I5686-01MS - 1	EFFLUENT-DAY1METALSMS	12192	3.8479	-245.1		PPB	13 Oct 2017 14:49:29	U	SMPL
I5686-01MSD - 1	EFFLUENT-DAY1METALSMSD	12031	3.7965	-245.1		PPB	13 Oct 2017 14:51:37	U	SMPL
I5867-01 - 1	101117-TIFR-E-U-B	-36	-0.0555	-245.1		PPB	13 Oct 2017 14:53:44	U	SMPL
CCV05 - 1	CCV05	15773	4.9910	-245.1		PPB	13 Oct 2017 14:55:51	U	SMPL
CCB05 - 1	CCB05	-30	-0.0536	-245.1		PPB	13 Oct 2017 15:00:28	U	SMPL
I5867-02 - 1	101117-TIFR-E-U-M	128	-0.0032	-245.1		PPB	13 Oct 2017 15:02:34	U	SMPL
I5867-03 - 1	101117-TIFR-E-U-S	122	-0.0051	-245.1		PPB	13 Oct 2017 15:04:41	U	SMPL
I5868-01 - 1	101117-TIDO-E-D-M	117	-0.0067	-245.1		PPB	13 Oct 2017 15:06:48	U	SMPL
I5868-02 - 1	101117-TIDO-E-U-M	130	-0.0025	-245.1		PPB	13 Oct 2017 15:08:55	U	SMPL
I5868-03 - 1	101117-TIFR-E-D-B	87	-0.0163	-245.1		PPB	13 Oct 2017 15:11:03	U	SMPL
I5868-04 - 1	101117-TIFR-E-D-M	96	-0.0134	-245.1		PPB	13 Oct 2017 15:13:10	U	SMPL
I5868-05 - 1	101117-TIFR-E-D-S	0	-0.0440	-245.1		PPB	13 Oct 2017 15:15:18	U	SMPL
I5870-01 - 1	101017-TIDO-E-D-M	20	-0.0376	-245.1		PPB	13 Oct 2017 15:17:27	U	SMPL
I5870-02 - 1	101017-TIDO-E-U-M	34	-0.0332	-245.1		PPB	13 Oct 2017 15:19:36	U	SMPL
CCV06 - 1	CCV06	15706	4.9696	-245.1		PPB	13 Oct 2017 15:21:43	U	SMPL
CCB06 - 1	CCB06	-73	-0.0673	-245.1		PPB	13 Oct 2017 15:23:49	U	SMPL
I5686-01LX5 - 1		134	-0.0012	-245.1		PPB	13 Oct 2017 15:25:57	U	SMPL
I5686-01A - 1		13657	4.3156	-245.1		PPB	13 Oct 2017 15:28:04	U	SMPL
CCV07 - 1	CCV07	15577	4.9285	-245.1		PPB	13 Oct 2017 15:30:11	U	SMPL
CCB07 - 1	CCB07	-32	-0.0542	-245.1		PPB	13 Oct 2017 15:32:18	U	SMPL