

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB90806

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP58767
calibration std. phosphate 0.5 ppm	WP58766
calibration std. phosphate 0.3 ppm	WP58765
calibration std. 0 ppm	WP58762
phosphate ICV std.	WP58769
phosphate CCV std.	WP58768
calibration std. phosphate 0.1 ppm	WP58764
calibration std. phosphate 0.05 ppm	WP58763
Combined reagent	WP58772
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036
Ascorbic Acid	WP58771

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	10/13/2017	11:36
2	CAL2	0.05	1	50	50	0.033	10/13/2017	11:36
3	CAL3	0.10	1	50	50	0.065	10/13/2017	11:37
4	CAL4	0.30	1	50	50	0.186	10/13/2017	11:37
5	CAL5	0.50	1	50	50	0.330	10/13/2017	11:38
6	CAL6	1.00	1	50	50	0.649	10/13/2017	11:38



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB90806

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.503	10/13/2017	11:39
2	ICB		1	50	50	0.000	0.001	10/13/2017	11:39
3	CCV1	0.50	1	50	50	0.328	0.506	10/13/2017	11:40
4	CCB1		1	50	50	0.000	0.001	10/13/2017	11:40
5	PB103260BL		1	50	50	0.000	0.001	10/13/2017	11:41
6	PB103260BS	0.50	1	50	50	0.325	0.501	10/13/2017	11:41
7	I5822-12		1	50	50	0.094	0.146	10/13/2017	11:42
8	I5822-12DUP		1	50	50	0.093	0.144	10/13/2017	11:42
9	I5822-12MS	0.50	1	50	50	0.421	0.649	10/13/2017	11:43
10	I5822-12MSD	0.50	1	50	50	0.420	0.647	10/13/2017	11:43
11	I5852-01		1	50	50	1.320	2.031	10/13/2017	11:44
12	I5852-01		10	50	50	0.313	0.483	10/13/2017	11:44
13	CCV2	0.50	1	50	50	0.329	0.507	10/13/2017	11:45
14	CCB2		1	50	50	0.000	0.001	10/13/2017	11:45