



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90827

Temp.: 17.2 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/14/2017	13:25
2	CAL2	1	Water			7.01	10/14/2017	13:26
3	CAL3	1	Water			10.03	10/14/2017	13:27
4	ICV	1	Water			7.00	10/14/2017	13:30
5	CCV1	1	Water			2.01	10/14/2017	13:32
6	I5905-03	1	Water			7.51	10/14/2017	13:33
7	I5905-03DUP	1	Water			7.52	10/14/2017	13:35
8	CCV2	1	Water			12.01	10/14/2017	13:37

1590827

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I5905-1 WorkList ID : 104203 Date : 10/14/2017 1:07:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/13/2017	Water	I5905-03	Corrosivity	Cool 4 deg C	SEAC01	N52	GROUNDWATER	10/13/2017	9040C

Date/Time 10.14.17 11:07 PM
Received by: JP
Relinquished by: CP

Date/Time 10-14-17 2:12 PM
Received by: CB
Relinquished by: JP