

7470A


LB90836

INSTRUMENT ID: LV1

Linear

 μ Abs.:

35756



A= 0.0000e+000

B= 2.8347e-004

C= 3.5306e-002

Rho= 0.9995057

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.034	0.034	-3	0.000	-3				
0.2	0.200	0.235	0.035	704	0.0 %	704				
2.5	2.500	2.554	0.054	8884	0.0 %	8884				
5.0	5.000	4.891	-0.109	17129	0.0 %	17129				
7.5	7.500	7.315	-0.185	25680	0.0 %	25680				
10.0	10.000	10.171	0.171	35756	0.0 %	35756				

LB90836 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Oct 2017 10:39:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.00007470A	PPB		16 Oct 2017 10:47:37	S	Std
0.2 - 1	80.2	704	-	0.20007470A	PPB		16 Oct 2017 10:49:46	S	Std
2.5 - 1	52.5	8884	-	2.50007470A	PPB		16 Oct 2017 10:51:55	S	Std
5.0 - 1	35	17129	-	5.00007470A	PPB		16 Oct 2017 10:54:03	S	Std
7.5 - 1	57.5	25680	-	7.50007470A	PPB		16 Oct 2017 10:56:10	S	Std
10.0 - 1	510	35756	-	10.00007470A	PPB		16 Oct 2017 10:58:18	S	Std
ICV10 - 1	ICV10	13469	3.8534	-7470A	PPB		16 Oct 2017 11:00:49	U	SMPL
ICB10 - 1	ICB10	-168	-0.0123	-7470A	PPB		16 Oct 2017 11:02:57	U	SMPL
CCV12 - 1	CCV12	17219	4.9164	-7470A	PPB		16 Oct 2017 11:05:03	U	SMPL
CCB12 - 1	CCB12	-164	-0.0112	-7470A	PPB		16 Oct 2017 11:07:10	U	SMPL
CRA - 1	CRA	679	0.2278	-7470A	PPB		16 Oct 2017 11:09:16	U	SMPL
HighStd - 1	HighStd	36633	10.4198	-7470A	PPB		16 Oct 2017 11:11:23	U	SMPL
ChkStd - 1	ChkStd	23513	6.7006	-7470A	PPB		16 Oct 2017 11:13:30	U	SMPL
PB103254BL - 1	PBW	-216	-0.0259	-7470A	PPB		16 Oct 2017 11:15:38	U	SMPL
PB103254BS - 1	LCSW	13897	3.9747	-7470A	PPB		16 Oct 2017 11:17:48	U	SMPL
I5837-01 - 1	20170685-FB	-171	-0.0132	-7470A	PPB		16 Oct 2017 11:19:55	U	SMPL
I5837-02 - 1	20170686-FB	-68	0.0160	-7470A	PPB		16 Oct 2017 11:22:03	U	SMPL
I5869-02 - 1	MW-5S-10-11-17	-19	0.0299	-7470A	PPB		16 Oct 2017 11:24:11	U	SMPL
I5869-03 - 1	MW-5D-10-11-17	-86	0.0109	-7470A	PPB		16 Oct 2017 11:26:20	U	SMPL
I5869-04 - 1	MW-6S-10-11-17	41026	11.6651	-7470A	PPB		16 Oct 2017 11:28:27	U	SMPL
CCV13 - 1	CCV13	16885	4.8218	-7470A	PPB		16 Oct 2017 11:30:33	U	SMPL
CCB13 - 1	CCB13	-187	-0.0177	-7470A	PPB		16 Oct 2017 11:32:42	U	SMPL
I5869-05 - 1	MW-6D-10-11-17	3	0.0362	-7470A	PPB		16 Oct 2017 11:34:49	U	SMPL
I5869-06 - 1	MW-7S-10-11-17	2860	0.8460	-7470A	PPB		16 Oct 2017 11:36:56	U	SMPL
I5869-07 - 1	MW-7D-10-11-17	7	0.0373	-7470A	PPB		16 Oct 2017 11:39:03	U	SMPL
I5869-08 - 1	DUP-01-10-11-17	1863	0.5634	-7470A	PPB		16 Oct 2017 11:41:10	U	SMPL
I5869-09 - 1	MW-08-10-11-17	285	0.1161	-7470A	PPB		16 Oct 2017 11:43:17	U	SMPL
I5869-10 - 1	MW-09-10-11-17	69	0.0549	-7470A	PPB		16 Oct 2017 11:45:24	U	SMPL
I5869-11 - 1	MW-10-10-11-17	128	0.0716	-7470A	PPB		16 Oct 2017 11:47:33	U	SMPL
I5869-11DUP - 1	MW-10-10-11-17DUP	107	0.0656	-7470A	PPB		16 Oct 2017 11:49:40	U	SMPL
I5869-12 - 1	I5869-11MS	14951	4.2735	-7470A	PPB		16 Oct 2017 11:51:49	U	SMPL
I5869-13 - 1	I5869-11MSD	15332	4.3815	-7470A	PPB		16 Oct 2017 11:54:41	U	SMPL
CCV14 - 1	CCV14	17463	4.9856	-7470A	PPB		16 Oct 2017 11:56:47	U	SMPL
CCB14 - 1	CCB14	-190	-0.0186	-7470A	PPB		16 Oct 2017 11:58:54	U	SMPL
I5869-15 - 1	MW-11-10-11-17	167	0.0826	-7470A	PPB		16 Oct 2017 12:01:01	U	SMPL
I5875-01 - 1	TK2064L-T1-T2-T3	935	0.3004	-7470A	PPB		16 Oct 2017 12:03:08	U	SMPL
I5905-03 - 1	GROUNDWATER	33	0.0447	-7470A	PPB		16 Oct 2017 12:05:15	U	SMPL
I5869-11LX5 - 1		8	0.0376	-7470A	PPB		16 Oct 2017 12:07:22	U	SMPL
I5869-11A - 1		14998	4.2868	-7470A	PPB		16 Oct 2017 12:09:30	U	SMPL
I5869-04DLX2 - 1		18720	5.3419	-7470A	PPB		16 Oct 2017 12:13:45	U	SMPL
CCV15 - 1	CCV15	17309	4.9419	-7470A	PPB		16 Oct 2017 12:15:55	U	SMPL
CCB15 - 1	CCB15	-173	-0.0137	-7470A	PPB		16 Oct 2017 12:18:05	U	SMPL