


7471A

LB 90837
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41615



A= 0.0000e+000

B= 2.4079e-004

C= 1.5144e-005

Rho= 0.9999195

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	0	0.000	0				
0.2	0.200	0.191	-0.009	792	0.0 %	792				
2.5	2.500	2.570	0.070	10674	0.0 %	10674				
5.0	5.000	4.912	-0.088	20401	0.0 %	20401				
7.5	7.500	7.506	0.006	31171	0.0 %	31171	0			
10.0	10.000	10.021	0.021	41615	0.0 %	41615				

LB90837 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 16 Oct 2017 12:32:08				
Sample ID	Extended ID	μ Abs.	Conc.	Strd ConcMethod	Units	Date	Type	Type
0.0 - 1	50	0	-	0.00007471A	PPB	16 Oct 2017 12:33:14	S	Std
0.2 - 1	50.2	792	-	0.20007471A	PPB	16 Oct 2017 12:35:32	S	Std
2.5 - 1	52.5	10674	-	2.50007471A	PPB	16 Oct 2017 12:37:50	S	Std
5.0 - 1	55	20401	-	5.00007471A	PPB	16 Oct 2017 12:40:08	S	Std
7.5 - 1	57.5	31171	-	7.50007471A	PPB	16 Oct 2017 12:42:25	S	Std
10.0 - 1	516	41615	-	10.00007471A	PPB	16 Oct 2017 12:44:42	S	Std
ICV11 - 1	ICV11	16313	3.9281	-7471A	PPB	16 Oct 2017 12:47:25	U	SMPL
ICB11 - 1	ICB11	-216	-0.0520	-7471A	PPB	16 Oct 2017 12:49:41	U	SMPL
CCV16 - 1	CCV16	20181	4.8595	-7471A	PPB	16 Oct 2017 12:51:58	U	SMPL
CCB16 - 1	CCB16	-192	-0.0462	-7471A	PPB	16 Oct 2017 12:54:13	U	SMPL
CRA - 1	CRA	781	0.1881	-7471A	PPB	16 Oct 2017 12:56:30	U	SMPL
HighStd - 1	HighStd	41747	10.0525	-7471A	PPB	16 Oct 2017 12:58:45	U	SMPL
ChkStd - 1	ChkStd	27345	6.5846	-7471A	PPB	16 Oct 2017 13:01:01	U	SMPL
PB103272BL - 1	PBS	-137	-0.0330	-7471A	PPB	16 Oct 2017 13:03:19	U	SMPL
PB103272BS - 1	LCSS	15552	3.7449	-7471A	PPB	16 Oct 2017 13:05:38	U	SMPL
I5865-01 - 1	CLEANSAND-10	184	0.0443	-7471A	PPB	16 Oct 2017 13:07:55	U	SMPL
I5865-01DUP - 1	CLEANSAND-10DUP	242	0.0583	-7471A	PPB	16 Oct 2017 13:10:12	U	SMPL
I5865-01MS - 1	CLEANSAND-10MS	15941	3.8385	-7471A	PPB	16 Oct 2017 13:12:30	U	SMPL
I5865-01MSD - 1	CLEANSAND-10MSD	15800	3.8046	-7471A	PPB	16 Oct 2017 13:14:48	U	SMPL
I5865-02 - 1	CLEANSAND-11	219	0.0527	-7471A	PPB	16 Oct 2017 13:17:04	U	SMPL
CCV17 - 1	CCV17	20155	4.8532	-7471A	PPB	16 Oct 2017 13:19:20	U	SMPL
CCB17 - 1	CCB17	-198	-0.0477	-7471A	PPB	16 Oct 2017 13:21:36	U	SMPL
I5865-03 - 1	CLEANSAND-12	184	0.0443	-7471A	PPB	16 Oct 2017 13:23:53	U	SMPL
I5890-16 - 1	ENGERT-COMP-1	313335	75.4495	-7471A	PPB	16 Oct 2017 13:26:08	U	SMPL
I5890-17 - 1	ENGERT-COMP-2	236178	56.8705	-7471A	PPB	16 Oct 2017 13:28:24	U	SMPL
I5895-01 - 1	SOIL-PILE	74565	17.9549	-7471A	PPB	16 Oct 2017 13:31:16	U	SMPL
I5896-01 - 1	B-19(0-10)COMP	100922	24.3015	-7471A	PPB	16 Oct 2017 13:34:04	U	SMPL
I5898-01 - 1	OR-02-101317	647	0.1558	-7471A	PPB	16 Oct 2017 13:36:24	U	SMPL
I5899-01 - 1	72-12020	129	0.0311	-7471A	PPB	16 Oct 2017 13:38:44	U	SMPL
I5901-01 - 1	S-1-20170688	1614	0.3887	-7471A	PPB	16 Oct 2017 13:41:01	U	SMPL
I5901-03 - 1	S-3-20170690	1255	0.3022	-7471A	PPB	16 Oct 2017 13:43:19	U	SMPL
I5901-05 - 1	S-5-20170692	1621	0.3903	-7471A	PPB	16 Oct 2017 13:45:35	U	SMPL
CCV18 - 1	CCV18	19334	4.6555	-7471A	PPB	16 Oct 2017 13:47:52	U	SMPL
CCB18 - 1	CCB18	-171	-0.0412	-7471A	PPB	16 Oct 2017 13:50:08	U	SMPL
I5901-06 - 1	S-6-20170693	3554	0.8558	-7471A	PPB	16 Oct 2017 13:52:25	U	SMPL
I5907-01 - 1	CONCRETE-FOUNDATION-COMP	182	0.0438	-7471A	PPB	16 Oct 2017 13:54:41	U	SMPL
I5908-01 - 1	T2-T3-COMP	3462	0.8336	-7471A	PPB	16 Oct 2017 13:56:58	U	SMPL
I5865-01LX5 - 1		3	0.0007	-7471A	PPB	16 Oct 2017 13:59:14	U	SMPL
I5865-01A - 1		15978	3.8474	-7471A	PPB	16 Oct 2017 14:01:31	U	SMPL
I5890-16DLX10 - 1		32426	7.8080	-7471A	PPB	16 Oct 2017 14:03:48	U	SMPL
I5890-17DLX10 - 1		23423	5.6402	-7471A	PPB	16 Oct 2017 14:06:06	U	SMPL
I5895-01DLX5 - 1		13924	3.3528	-7471A	PPB	16 Oct 2017 14:08:25	U	SMPL
I5896-01DLX5 - 1		19671	4.7367	-7471A	PPB	16 Oct 2017 14:10:43	U	SMPL
CCV19 - 1	CCV19	19830	4.7750	-7471A	PPB	16 Oct 2017 14:13:00	U	SMPL
CCB19 - 1	CCB19	-234	-0.0563	-7471A	PPB	16 Oct 2017 14:15:17	U	SMPL