



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90859

Temp.: 21.2 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/16/2017	17:10
2	CAL2	1	Water			7.01	10/16/2017	17:11
3	CAL3	1	Water			10.02	10/16/2017	17:13
4	ICV	1	Water			6.99	10/16/2017	17:15
5	CCV1	1	Water			2.01	10/16/2017	17:16
6	I5929-04	1	Solid	20.02	20	6.50	10/16/2017	17:18
7	I5929-08	1	Solid	20.04	20	6.02	10/16/2017	17:19
8	I5929-12	1	Solid	20.01	20	7.20	10/16/2017	17:21
9	I5931-03	1	Solid	20.04	20	6.03	10/16/2017	17:22
10	I5935-03	1	Solid	20.03	20	7.60	10/16/2017	17:24
11	I5935-06	1	Solid	20.03	20	7.75	10/16/2017	17:25
12	I5935-09	1	Solid	20.02	20	7.48	10/16/2017	17:26
13	I5935-12	1	Solid	20.04	20	7.87	10/16/2017	17:30
14	I5935-15	1	Solid	20.03	20	7.73	10/16/2017	17:31
15	I5935-15DUP	1	Solid	20.04	20	7.74	10/16/2017	17:32
16	CCV2	1	Water			12.02	10/16/2017	17:34

1300859

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5929 WorkList ID : 104269 Date : 10/16/2017 3:53:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/22/2017	Solid	I5929-04	Corrosivity	Cool 4 deg C	PSEG01	M52	TP-1-COMP	10/16/2017	9045C
10/22/2017	Solid	I5929-08	Corrosivity	Cool 4 deg C	PSEG01	M52	TP-2-COMP	10/16/2017	9045C
10/22/2017	Solid	I5929-12	Corrosivity	Cool 4 deg C	PSEG01	M52	TP-1-2-COMP	10/16/2017	9045C
10/19/2017	Solid	I5931-03	Corrosivity	Cool 4 deg C	PSEG05	M52	BU-01-101617	10/16/2017	9045C
10/19/2017	Solid	I5935-03	Corrosivity	Cool 4 deg C	PSEG05	M52	CL-02-101617-A	10/16/2017	9045C
10/19/2017	Solid	I5935-06	Corrosivity	Cool 4 deg C	PSEG05	M52	CL-02-101617-B	10/16/2017	9045C
10/19/2017	Solid	I5935-09	Corrosivity	Cool 4 deg C	PSEG05	M52	CL-02-101617-C	10/16/2017	9045C
10/19/2017	Solid	I5935-12	Corrosivity	Cool 4 deg C	PSEG05	M52	CL-02-101617-D	10/16/2017	9045C
10/19/2017	Solid	I5935-15	Corrosivity	Cool 4 deg C	PSEG05	M52	CL-02-101617-E	10/16/2017	9045C

Date/Time 10/16/17 4:00PM
Received by: JD
Relinquished by: CR

Date/Time 10/16/17 5:30PM
Received by: CR
Relinquished by: JD