



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90869

Temp.: 20.1 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	10/17/2017	11:45
2	CAL2	1	Water			7.01	10/17/2017	11:46
3	CAL3	1	Water			10.03	10/17/2017	11:47
4	ICV	1	Water			6.99	10/17/2017	11:48
5	CCV1	1	Water			2.02	10/17/2017	11:50
6	I5936-01	1	Water			7.29	10/17/2017	11:51
7	I5936-01DUP	1	Water			7.30	10/17/2017	11:52
8	CCV2	1	Water			12.01	10/17/2017	11:55

2B00869

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5936

WorkList ID : 104306

Date : 10/17/2017 11:32:31 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/16/2017	Water	I5936-01-B	pH	Cool 4 deg C	SEAC01	M31	GROUNDWATER-432-B	10/16/2017	SM 4500-PH B

Date/Time 10/17/17 11:32 AM
 Received by: 711
 Relinquished by: CD

Date/Time 10/17/17 12:30 PM
 Received by: CD
 Relinquished by: JH