


ISM02.4

LB 90871
INSTRUMENT ID: CV2

Linear

 μ Abs.:

40635



A= 0.0000e+000

B= 2.4534e-004

C= -2.2845e-002

Rho= 0.9999553

Accept= Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.021	-0.021	9	0.000	9				
0.2	0.200	0.183	-0.017	841	0.0 %	841				
2.5	2.500	2.505	0.005	10304	0.0 %	10304				
5.0	5.000	5.049	0.049	20674	0.0 %	20674				
7.5	7.500	7.536	0.036	30809	0.0 %	30809				
10.0	10.000	9.947	-0.053	40635	0.0 %	40635				

90D
-9
0
0
-1
NB
10/17/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	9	-		0 ISM02.4	PPB	10/17/2017 12:13	Std
	0.2 S0.2	841	-		0.2 ISM02.4	PPB	10/17/2017 12:15	Std
	2.5 S2.5	10304	-		2.5 ISM02.4	PPB	10/17/2017 12:19	Std
	5 S5	20674	-		5 ISM02.4	PPB	10/17/2017 12:23	Std
	7.5 S7.5	30809	-		7.5 ISM02.4	PPB	10/17/2017 12:25	Std
	10 S10	40635	-		10 ISM02.4	PPB	10/17/2017 12:27	Std
ICV	ICV	15204	3.7074	-	ISM02.4	PPB	10/17/2017 12:30	SMPL
ICB	ICB	-173	-0.0653	-	ISM02.4	PPB	10/17/2017 12:32	SMPL
CCV029	CCV029	20675	5.0496	-	ISM02.4	PPB	10/17/2017 12:35	SMPL
CCB029	CCB029	-182	-0.0675	-	ISM02.4	PPB	10/17/2017 12:37	SMPL
PB103285BL	PBW003	12	-0.0199	-	ISM02.4	PPB	10/17/2017 12:39	SMPL
I5742-01	MJJ4L0	-32	-0.0307	-	ISM02.4	PPB	10/17/2017 12:41	SMPL
I5742-02	MJJ4L1	4	-0.0219	-	ISM02.4	PPB	10/17/2017 12:44	SMPL
I5742-03	MJJ4L2	-14	-0.0263	-	ISM02.4	PPB	10/17/2017 12:46	SMPL
I5742-04	MJJ4L3	-55	-0.0363	-	ISM02.4	PPB	10/17/2017 12:48	SMPL
I5742-05	MJJ4L4	-15	-0.0265	-	ISM02.4	PPB	10/17/2017 12:50	SMPL
I5742-06	MJJ4L5	-42	-0.0331	-	ISM02.4	PPB	10/17/2017 12:53	SMPL
I5742-07	MJJ4L6	39	-0.0133	-	ISM02.4	PPB	10/17/2017 12:55	SMPL
I5742-08	MJJ4L6D	-19	-0.0275	-	ISM02.4	PPB	10/17/2017 12:57	SMPL
I5742-09	MJJ4L6S	4089	0.9804	-	ISM02.4	PPB	10/17/2017 13:00	SMPL
I5742-10	MJJ4L7	-53	-0.0358	-	ISM02.4	PPB	10/17/2017 13:02	SMPL
I5742-11	MJJ4L8	-38	-0.0322	-	ISM02.4	PPB	10/17/2017 13:04	SMPL
I5742-12	MJJ4L9	14	-0.0194	-	ISM02.4	PPB	10/17/2017 13:06	SMPL
I5742-13	MJJ4M0	-17	-0.027	-	ISM02.4	PPB	10/17/2017 13:09	SMPL
I5742-14	MJJ4M1	1	-0.0226	-	ISM02.4	PPB	10/17/2017 13:11	SMPL
I5742-15	MJJ4M2	-37	-0.0319	-	ISM02.4	PPB	10/17/2017 13:13	SMPL
I5742-16	MJJ4M3	-33	-0.0309	-	ISM02.4	PPB	10/17/2017 13:16	SMPL
PB103286BL	PBW003	18	-0.0184	-	ISM02.4	PPB	10/17/2017 13:18	SMPL
I5743-01	MJJ4N2	6	-0.0214	-	ISM02.4	PPB	10/17/2017 13:20	SMPL
I5743-02	MJJ4N3	-28	-0.0297	-	ISM02.4	PPB	10/17/2017 13:22	SMPL
CCV030	CCV030	20782	5.0759	-	ISM02.4	PPB	10/17/2017 13:25	SMPL
CCB030	CCB030	-187	-0.0687	-	ISM02.4	PPB	10/17/2017 13:27	SMPL
I5743-03	MJJ4N4	-20	-0.0278	-	ISM02.4	PPB	10/17/2017 13:29	SMPL
I5743-04	MJJ4N4D	-47	-0.0344	-	ISM02.4	PPB	10/17/2017 13:31	SMPL
I5743-05	MJJ4N4S	3937	0.9431	-	ISM02.4	PPB	10/17/2017 13:34	SMPL
I5743-06	MJJ4N5	-4	-0.0238	-	ISM02.4	PPB	10/17/2017 13:36	SMPL
I5743-07	MJJ4N6	-50	-0.0351	-	ISM02.4	PPB	10/17/2017 13:38	SMPL
I5743-08	MJJ4N7	-32	-0.0307	-	ISM02.4	PPB	10/17/2017 13:41	SMPL
PB103292BL	PBW003	5	-0.0216	-	ISM02.4	PPB	10/17/2017 13:43	SMPL
I5826-11	EQB-3-101117	73	-0.0049	-	ISM02.4	PPB	10/17/2017 13:45	SMPL
I5829-01	TWP-B-38	564	0.1155	-	ISM02.4	PPB	10/17/2017 13:47	SMPL
I5829-02	TWP-B-63	467	0.0917	-	ISM02.4	PPB	10/17/2017 13:50	SMPL
I5871-01	TWP-B-61	385	0.0716	-	ISM02.4	PPB	10/17/2017 13:52	SMPL
I5871-02	TWP-B-13	70	-0.0057	-	ISM02.4	PPB	10/17/2017 13:54	SMPL
I5902-08	TWP-B-18	101	0.0019	-	ISM02.4	PPB	10/17/2017 13:57	SMPL
CCV031	CCV031	20796	5.0793	-	ISM02.4	PPB	10/17/2017 13:59	SMPL

LB90871

INSTRUMENT ID : CV1

CCB031

CCB031

-177 -0.0663 -

ISM02.4 PPB

10/17/2017 14:01 SMPL