



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB90876

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	10/17/2017	09:10
2	I5875-02		1	>100.0	>212.0	10/17/2017	19:42
3	I5875-02DUP		1	>100.0	>212.0	10/17/2017	10:15
4	I5905-03		1	>100.0	>212.0	10/17/2017	10:46

Result = (Celsius * 1.8) + 32

LB 90876

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP101717

WorkList ID : 104313

Date : 10/17/2017 08:30 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/17/2017	Water	I5875-02	Flash Point	Cool 4 deg C	PSEG01	H52	TK2064L-T1-T2-T3	10/12/2017	1010A
10/27/2017	Water	I5905-03	Flash Point	Cool 4 deg C	SEAC01	N52	GROUNDWATER	10/13/2017	1010A

Date/Time 10/17/17 8:40 AM
Received by: KS
Relinquished by: CP

Date/Time 10/17/17 10:50 AM
Received by: CP
Relinquished by: KS