


7471A

LB90886
INSTRUMENT ID: CV2

Linear

 μ Abs.:

43223



A= 0.0000e+000

B= 2.3444e-004

C= 9.9820e-003

Rho= 0.9994529

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.033	-0.033	-185	0.000	-185				
0.2	0.200	0.223	0.023	909	0.0 %	909				
2.5	2.500	2.646	0.146	11243	0.0 %	11243				
5.0	5.000	4.920	-0.080	20944	0.0 %	20944				
7.5	7.500	7.301	-0.199	31100	0.0 %	31100	0			
10.0	10.000	10.143	0.143	43223	0.0 %	43223				

LB90886 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 17 Oct 2017 17:28:16

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-185	-	0.00007471A	PPB		17 Oct 2017 17:57:12	S	Std
0.2 - 1	50.2	909	-	0.20007471A	PPB		17 Oct 2017 17:59:30	S	Std
2.5 - 1	52.5	11243	-	2.50007471A	PPB		17 Oct 2017 18:01:47	S	Std
5.0 - 1	55	20944	-	5.00007471A	PPB		17 Oct 2017 18:04:05	S	Std
7.5 - 1	57.5	31100	-	7.50007471A	PPB		17 Oct 2017 18:06:22	S	Std
10.0 - 1	510	43223	-	10.00007471A	PPB		17 Oct 2017 18:16:47	S	Std
ICV14 - 1	ICV14	16567	3.8940	-7471A	PPB		17 Oct 2017 18:19:29	U	SMPL
ICB14 - 1	ICB14	191	0.0548	-7471A	PPB		17 Oct 2017 18:21:45	U	SMPL
CCV35 - 1	CCV35	21241	4.9898	-7471A	PPB		17 Oct 2017 18:24:02	U	SMPL
CCB35 - 1	CCB35	-85	-0.0099	-7471A	PPB		17 Oct 2017 18:26:17	U	SMPL
CRA - 1	CRA	945	0.2315	-7471A	PPB		17 Oct 2017 18:28:34	U	SMPL
HighStd - 1	HighStd	43586	10.2284	-7471A	PPB		17 Oct 2017 18:30:50	U	SMPL
ChkStd - 1	ChkStd	29370	6.8955	-7471A	PPB		17 Oct 2017 18:33:05	U	SMPL
PB103359BL - 1	PBS	82	0.0292	-7471A	PPB		17 Oct 2017 18:35:24	U	SMPL
PB103359BS - 1	LCSS	15640	3.6766	-7471A	PPB		17 Oct 2017 18:37:43	U	SMPL
I5879-07 - 1	TP-1(6-8)	4895	1.1576	-7471A	PPB		17 Oct 2017 18:40:00	U	SMPL
I5924-03 - 1	BACKFILL-COMP-1	396	0.1028	-7471A	PPB		17 Oct 2017 18:42:17	U	SMPL
I5929-01 - 1	TP-1-COMP	904	0.2219	-7471A	PPB		17 Oct 2017 18:44:35	U	SMPL
I5929-05 - 1	TP-2-COMP	1099	0.2676	-7471A	PPB		17 Oct 2017 18:46:53	U	SMPL
I5929-09 - 1	TP-1-2-COMP	1910	0.4578	-7471A	PPB		17 Oct 2017 18:49:09	U	SMPL
CCV36 - 1	CCV36	21149	4.9682	-7471A	PPB		17 Oct 2017 18:51:25	U	SMPL
CCB36 - 1	CCB36	-76	-0.0078	-7471A	PPB		17 Oct 2017 18:53:41	U	SMPL
I5930-01 - 1	FK-GF-02-012-A	1851	0.4439	-7471A	PPB		17 Oct 2017 18:55:58	U	SMPL
I5930-01DUP - 1	FK-GF-02-012-ADUP	1790	0.4296	-7471A	PPB		17 Oct 2017 18:58:13	U	SMPL
I5930-01MS - 1	FK-GF-02-012-AMS	17314	4.0691	-7471A	PPB		17 Oct 2017 19:00:30	U	SMPL
I5930-01MSD - 1	FK-GF-02-012-AMSD	16707	3.9268	-7471A	PPB		17 Oct 2017 19:02:46	U	SMPL
I5930-03 - 1	FK-GF-02-012-B	1818	0.4362	-7471A	PPB		17 Oct 2017 19:05:03	U	SMPL
I5931-01 - 1	BU-01-101617	1271	0.3080	-7471A	PPB		17 Oct 2017 19:07:20	U	SMPL
I5935-01 - 1	CL-02-101617-A	3705	0.8786	-7471A	PPB		17 Oct 2017 19:09:38	U	SMPL
I5935-04 - 1	CL-02-101617-B	2561	0.6104	-7471A	PPB		17 Oct 2017 19:11:55	U	SMPL
I5935-07 - 1	CL-02-101617-C	3841	0.9105	-7471A	PPB		17 Oct 2017 19:14:13	U	SMPL
I5935-10 - 1	CL-02-101617-D	2918	0.6941	-7471A	PPB		17 Oct 2017 19:16:29	U	SMPL
CCV37 - 1	CCV37	21329	5.0104	-7471A	PPB		17 Oct 2017 19:18:45	U	SMPL
CCB37 - 1	CCB37	-36	0.0015	-7471A	PPB		17 Oct 2017 19:21:02	U	SMPL
I5935-13 - 1	CL-02-101617-E	1263	0.3061	-7471A	PPB		17 Oct 2017 19:23:19	U	SMPL
PB103360BL - 1	PBS	179	0.0519	-7471A	PPB		17 Oct 2017 19:25:35	U	SMPL
PB103360BS - 1	LCSS	14399	3.3857	-7471A	PPB		17 Oct 2017 19:27:52	U	SMPL
I5922-01 - 1	ENV-115-7(0-5)	1601	0.3853	-7471A	PPB		17 Oct 2017 19:30:08	U	SMPL
I5922-02 - 1	ENV-115-6(0-5)	2133	0.5100	-7471A	PPB		17 Oct 2017 19:32:25	U	SMPL
I5922-02DUP - 1	ENV-115-6(0-5)DUP	2050	0.4906	-7471A	PPB		17 Oct 2017 19:34:42	U	SMPL
I5922-02MS - 1	ENV-115-6(0-5)MS	17804	4.1840	-7471A	PPB		17 Oct 2017 19:36:59	U	SMPL
I5922-02MSD - 1	ENV-115-6(0-5)MSD	18890	4.4386	-7471A	PPB		17 Oct 2017 19:39:17	U	SMPL
I5922-03 - 1	ENV-125-11(30-40)	143	0.0435	-7471A	PPB		17 Oct 2017 19:41:35	U	SMPL
I5922-04 - 1	ENV-125-11(40-50)	240	0.0662	-7471A	PPB		17 Oct 2017 19:43:52	U	SMPL
CCV38 - 1	CCV38	21395	5.0259	-7471A	PPB		17 Oct 2017 19:46:08	U	SMPL
CCB38 - 1	CCB38	19	0.0144	-7471A	PPB		17 Oct 2017 19:48:25	U	SMPL
I5922-05 - 1	ENV-125-11(50-58)	281	0.0759	-7471A	PPB		17 Oct 2017 19:50:42	U	SMPL
I5922-06 - 1	ENV-115-1(0-5)	315	0.0838	-7471A	PPB		17 Oct 2017 19:52:58	U	SMPL
I5922-07 - 1	ENV-115-1(5-10)	437	0.1124	-7471A	PPB		17 Oct 2017 19:55:15	U	SMPL
I5922-08 - 1	ENV-115-1(10-15)	598	0.1502	-7471A	PPB		17 Oct 2017 19:57:31	U	SMPL
I5922-09 - 1	EPE-M6(9-11)	346	0.0911	-7471A	PPB		17 Oct 2017 19:59:48	U	SMPL
I5930-01LX5 - 1		494	0.1258	-7471A	PPB		17 Oct 2017 20:02:05	U	SMPL
I5930-01A - 1		18086	4.2501	-7471A	PPB		17 Oct 2017 20:04:22	U	SMPL
I5922-02LX5 - 1		391	0.1016	-7471A	PPB		17 Oct 2017 20:06:40	U	SMPL
I5922-02A - 1		18491	4.3450	-7471A	PPB		17 Oct 2017 20:08:58	U	SMPL
CCV39 - 1	CCV39	21444	5.0373	-7471A	PPB		17 Oct 2017 20:11:14	U	SMPL
CCB39 - 1	CCB39	28	0.0165	-7471A	PPB		17 Oct 2017 20:13:30	U	SMPL