



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90891

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/17/2017	16:15
2	CAL2	1	Water			7.00	10/17/2017	16:17
3	CAL3	1	Water			10.04	10/17/2017	16:19
4	ICV	1	Water			7.00	10/17/2017	16:21
5	CCV1	1	Water			2.01	10/17/2017	16:22
6	I5943-03	1	Solid	20.02	20	8.74	10/17/2017	16:23
7	I5951-03	1	Solid	20.04	20	8.86	10/17/2017	16:25
8	I5952-03	1	Solid	20.05	20	5.67	10/17/2017	16:26
9	I5953-03	1	Solid	20.02	20	8.43	10/17/2017	16:30
10	I5953-06	1	Solid	20.03	20	9.05	10/17/2017	16:31
11	I5953-06DUP	1	Solid	20.04	20	9.06	10/17/2017	16:33
12	CCV2	1	Water			12.02	10/17/2017	16:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5944

WorkList ID : 104317

Date : 10/17/2017 3:22:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/20/2017	Solid	I5943-03	Corrosivity	Cool 4 deg C	PSEG05	M52	SU-03-101717	10/17/2017	9045C
10/20/2017	Solid	I5951-03	Corrosivity	Cool 4 deg C	PSEG01	M52	294-COMP	10/17/2017	9045C
10/20/2017	Solid	I5952-03	Corrosivity	Cool 4 deg C	PSEG01	M52	RO-COMP	10/17/2017	9045C
10/20/2017	Solid	I5953-03	Corrosivity	Cool 4 deg C	PSEG05	M52	EO-02-101717-A	10/17/2017	9045C
10/20/2017	Solid	I5953-06	Corrosivity	Cool 4 deg C	PSEG05	M52	EO-02-101717-B	10/17/2017	9045C

Date/Time 10-17-17 3:22 PM
 Received by: JP
 Relinquished by: JP

Date/Time 10-17-17 6:25 PM
 Received by: JP
 Relinquished by: JP