

LB 90896


7471A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

42300



A= 0.0000e+000

B= 2.3794e-004

C= -4.3533e-002

Rho= 0.9999174

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.041	-0.041	11	0.000	11				
0.2	0.200	0.257	0.057	1262	0.0 %	1262				
2.5	2.500	2.533	0.033	10830	0.0 %	10830				
5.0	5.000	4.917	-0.083	20848	0.0 %	20848				
7.5	7.500	7.512	0.012	31754	0.0 %	31754	0			
10.0	10.000	10.021	0.021	42300	0.0 %	42300				

LB90896 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 18 Oct 2017 09:05:50				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	SO	11	-	0.0000	7471A	PPB	18 Oct 2017 09:12:06	S Std
0.2 - 1	SO-2	1262	-	0.2000	7471A	PPB	18 Oct 2017 09:14:24	S Std
2.5 - 1	SO-5	10830	-	2.5000	7471A	PPB	18 Oct 2017 09:16:42	S Std
5.0 - 1	SO	20848	-	5.0000	7471A	PPB	18 Oct 2017 09:18:59	S Std
7.5 - 1	SO-5	31754	-	7.5000	7471A	PPB	18 Oct 2017 09:21:16	S Std
10.0 - 1	SO	42300	-	10.0000	7471A	PPB	18 Oct 2017 09:23:33	S Std
ICV15 - 1	ICV15	16204	3.8121	-7471A	PPB	PPB	18 Oct 2017 09:26:19	U SMPL
ICB15 - 1	ICB15	-163	-0.0823	-7471A	PPB	PPB	18 Oct 2017 09:28:35	U SMPL
CCV40 - 1	CCV40	21088	4.9742	-7471A	PPB	PPB	18 Oct 2017 09:30:51	U SMPL
CCB40 - 1	CCB40	-171	-0.0842	-7471A	PPB	PPB	18 Oct 2017 09:33:07	U SMPL
CRA - 1	CRA	1249	0.2537	-7471A	PPB	PPB	18 Oct 2017 09:35:22	U SMPL
HighStd - 1	HighStd	42393	10.0436	-7471A	PPB	PPB	18 Oct 2017 09:37:38	U SMPL
ChkStd - 1	ChkStd	29265	6.9199	-7471A	PPB	PPB	18 Oct 2017 09:39:54	U SMPL
PB103392BL - 1	PBS	-30	-0.0507	-7471A	PPB	PPB	18 Oct 2017 09:44:44	U SMPL
PB103392BS - 1	LCSS	15820	3.7207	-7471A	PPB	PPB	18 Oct 2017 09:47:00	U SMPL
I5942-01 - 1	SOIL-1-2	6244	1.4422	-7471A	PPB	PPB	18 Oct 2017 09:49:18	U SMPL
I5942-01DUP - 1	SOIL-1-2DUP	6482	1.4988	-7471A	PPB	PPB	18 Oct 2017 09:51:35	U SMPL
I5942-01MS - 1	SOIL-1-2MS	21408	5.0504	-7471A	PPB	PPB	18 Oct 2017 09:53:53	U SMPL
I5942-01MSD - 1	SOIL-1-2MSD	22483	5.3062	-7471A	PPB	PPB	18 Oct 2017 09:56:10	U SMPL
I5943-01 - 1	SU-03-101717	1678	0.3557	-7471A	PPB	PPB	18 Oct 2017 09:58:27	U SMPL
CCV41 - 1	CCV41	21233	5.0087	-7471A	PPB	PPB	18 Oct 2017 10:00:43	U SMPL
CCB41 - 1	CCB41	-159	-0.0814	-7471A	PPB	PPB	18 Oct 2017 10:02:59	U SMPL
I5944-01 - 1	CONCRETE-PILE	2466	0.5432	-7471A	PPB	PPB	18 Oct 2017 10:05:16	U SMPL
I5951-01 - 1	294-COMP	13389	3.1423	-7471A	PPB	PPB	18 Oct 2017 10:07:32	U SMPL
I5952-01 - 1	RO-COMP	4872	1.1157	-7471A	PPB	PPB	18 Oct 2017 10:09:48	U SMPL
I5953-01 - 1	EO-02-101717-A	2428	0.5342	-7471A	PPB	PPB	18 Oct 2017 10:12:04	U SMPL
I5953-04 - 1	EO-02-101717-B	2424	0.5332	-7471A	PPB	PPB	18 Oct 2017 10:14:21	U SMPL
I5956-01 - 1	EPE-M-6(50-52)	464	0.0669	-7471A	PPB	PPB	18 Oct 2017 10:16:38	U SMPL
I5956-02 - 1	EPE-M-6(70-72)	647	0.1104	-7471A	PPB	PPB	18 Oct 2017 10:18:55	U SMPL
I5956-03 - 1	EPE-120-1(0-5)	1222	0.2472	-7471A	PPB	PPB	18 Oct 2017 10:21:12	U SMPL
I5956-04 - 1	EPE-120-1(5-10)	556	0.0888	-7471A	PPB	PPB	18 Oct 2017 10:23:30	U SMPL
I5956-05 - 1	EPE-120-1(10-12)	752	0.1354	-7471A	PPB	PPB	18 Oct 2017 10:25:47	U SMPL
CCV42 - 1	CCV42	21159	4.9911	-7471A	PPB	PPB	18 Oct 2017 10:28:03	U SMPL
CCB42 - 1	CCB42	-129	-0.0742	-7471A	PPB	PPB	18 Oct 2017 10:30:20	U SMPL
I5956-06 - 1	EPE-120-1(15-17)	422	0.0569	-7471A	PPB	PPB	18 Oct 2017 10:32:37	U SMPL
I5956-07 - 1	EPE-120-1(20-22)	343	0.0381	-7471A	PPB	PPB	18 Oct 2017 10:34:54	U SMPL
I5956-08 - 1	EPE-120-1(25-27)	134	-0.0116	-7471A	PPB	PPB	18 Oct 2017 10:37:10	U SMPL
I5956-09 - 1	EPE-108-2(10-12)	255	0.0171	-7471A	PPB	PPB	18 Oct 2017 10:39:26	U SMPL
I5956-10 - 1	EPE-108-2(20-22)	191	0.0019	-7471A	PPB	PPB	18 Oct 2017 10:41:43	U SMPL
I5942-01LX5 - 1		1147	0.2294	-7471A	PPB	PPB	18 Oct 2017 10:44:00	U SMPL
I5942-01A - 1		23281	5.4960	-7471A	PPB	PPB	18 Oct 2017 10:46:17	U SMPL
CCV43 - 1	CCV43	21078	4.9719	-7471A	PPB	PPB	18 Oct 2017 10:48:35	U SMPL
CCB43 - 1	CCB43	-66	-0.0592	-7471A	PPB	PPB	18 Oct 2017 10:50:54	U SMPL