


245.1

LB90918
INSTRUMENT ID: CV1

Linear

 μ Abs.:

32022



A= 0.0000e+000

B= 3.1276e-004

C= -2.9715e-002

Rho= 0.9999953

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	117	0.000	117				
0.2	0.200	0.186	-0.014	691	0.0 %	691				
2.5	2.500	2.497	-0.003	8078	0.0 %	8078				
5.0	5.000	5.011	0.011	16118	0.0 %	16118				
7.5	7.500	7.513	0.013	24118	0.0 %	24118				
10.0	10.000	9.985	-0.015	32022	0.0 %	32022				

LB90918 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 18 Oct 2017 16:13:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	117	-	0.0000	245.1	PPB	18 Oct 2017 16:21:31	S	Std
0.2 - 1	50.2	691	-	0.2000	245.1	PPB	18 Oct 2017 16:23:39	S	Std
2.5 - 1	52.5	8078	-	2.5000	245.1	PPB	18 Oct 2017 16:25:48	S	Std
5.0 - 1	55	16118	-	5.0000	245.1	PPB	18 Oct 2017 16:27:56	S	Std
7.5 - 1	57.5	24118	-	7.5000	245.1	PPB	18 Oct 2017 16:32:35	S	Std
10.0 - 1	510	32022	-	10.0000	245.1	PPB	18 Oct 2017 16:34:42	S	Std
ICV17 - 1	ICV17	12471	3.8707	-245.1		PPB	18 Oct 2017 16:37:32	U	SMPL
ICB17 - 1	ICB17	-52	-0.0460	-245.1		PPB	18 Oct 2017 16:39:38	U	SMPL
CCV48 - 1	CCV48	16213	5.0410	-245.1		PPB	18 Oct 2017 16:41:45	U	SMPL
CCB48 - 1	CCB48	-64	-0.0497	-245.1		PPB	18 Oct 2017 16:43:52	U	SMPL
CRA - 1	CRA	637	0.1695	-245.1		PPB	18 Oct 2017 16:46:00	U	SMPL
HighStd - 1	HighStd	32283	10.0670	-245.1		PPB	18 Oct 2017 16:48:06	U	SMPL
ChkStd - 1	ChkStd	22462	6.9954	-245.1		PPB	18 Oct 2017 16:50:13	U	SMPL
PB103389BL - 1	PBW	-147	-0.0757	-245.1		PPB	18 Oct 2017 16:52:22	U	SMPL
PB103389BS - 1	LCSW	13594	4.2219	-245.1		PPB	18 Oct 2017 16:54:31	U	SMPL
I5927-01 - 1	101317-TIDO-E-D-M	-59	-0.0482	-245.1		PPB	18 Oct 2017 16:56:39	U	SMPL
I5927-01DUP - 1	101317-TIDO-E-D-MDUP	34	-0.0191	-245.1		PPB	18 Oct 2017 16:58:47	U	SMPL
I5927-01MS - 1	101317-TIDO-E-D-MMS	14839	4.6113	-245.1		PPB	18 Oct 2017 17:00:55	U	SMPL
I5927-01MSD - 1	101317-TIDO-E-D-MMSD	14761	4.5869	-245.1		PPB	18 Oct 2017 17:03:03	U	SMPL
I5927-02 - 1	101317-TIDO-E-U-M	-41	-0.0425	-245.1		PPB	18 Oct 2017 17:05:10	U	SMPL
CCV49 - 1	CCV49	16034	4.9850	-245.1		PPB	18 Oct 2017 17:07:17	U	SMPL
CCB49 - 1	CCB49	-45	-0.0438	-245.1		PPB	18 Oct 2017 17:09:24	U	SMPL
I5927-03 - 1	101317-TIFR-E-D-B	45	-0.0156	-245.1		PPB	18 Oct 2017 17:11:32	U	SMPL
I5927-04 - 1	101317-TIFR-E-D-S	56	-0.0122	-245.1		PPB	18 Oct 2017 17:13:39	U	SMPL
I5927-07 - 1	101317-TIFR-E-U-B	139	0.0138	-245.1		PPB	18 Oct 2017 17:15:45	U	SMPL
I5927-08 - 1	101317-TIFR-E-U-S	85	-0.0031	-245.1		PPB	18 Oct 2017 17:17:52	U	SMPL
I5927-01LX5 - 1		76	-0.0059	-245.1		PPB	18 Oct 2017 17:20:00	U	SMPL
I5927-01A - 1		15660	4.8680	-245.1		PPB	18 Oct 2017 17:22:07	U	SMPL
CCV50 - 1	CCV50	15941	4.9559	-245.1		PPB	18 Oct 2017 17:24:15	U	SMPL
CCB50 - 1	CCB50	-55	-0.0469	-245.1		PPB	18 Oct 2017 17:26:23	U	SMPL