


7470A

LB 90931
INSTRUMENT ID: CV2

Linear

 μ Abs.:

33598



A= 0.0000e+000

B= 2.9516e-004

C= -5.3264e-003

Rho= 0.9998409

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	4	0.000	4				
0.2	0.200	0.202	0.002	703	0.0 %	703				
2.5	2.500	2.467	-0.033	8378	0.0 %	8378				
5.0	5.000	4.993	-0.007	16934	0.0 %	16934				
7.5	7.500	7.630	0.130	25870	0.0 %	25870				
10.0	10.000	9.911	-0.089	33598	0.0 %	33598				

LB90931 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Oct 2017 09:33:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	4	-	0.0000	7470A	PPB	19 Oct 2017 09:37:30	S	Std
0.2 - 1	50.2	703	-	0.2000	7470A	PPB	19 Oct 2017 09:39:39	S	Std
2.5 - 1	52.5	8378	-	2.5000	7470A	PPB	19 Oct 2017 09:41:47	S	Std
5.0 - 1	55	16934	-	5.0000	7470A	PPB	19 Oct 2017 09:43:55	S	Std
7.5 - 1	57.5	25870	-	7.5000	7470A	PPB	19 Oct 2017 09:46:03	S	Std
10.0 - 1	510	33598	-	10.0000	7470A	PPB	19 Oct 2017 09:48:11	S	Std
ICV18 - 1	ICV18	13535	3.9896	-	7470A	PPB	19 Oct 2017 09:50:45	U	SMPL
ICB18 - 1	ICB18	-143	-0.0475	-	7470A	PPB	19 Oct 2017 09:52:52	U	SMPL
CCV51 - 1	CCV51	17303	5.1017	-	7470A	PPB	19 Oct 2017 09:54:59	U	SMPL
CCB51 - 1	CCB51	-149	-0.0493	-	7470A	PPB	19 Oct 2017 09:57:05	U	SMPL
CRA - 1	CRA	701	0.2016	-	7470A	PPB	19 Oct 2017 09:59:13	U	SMPL
HighStd - 1	HighStd	33956	10.0170	-	7470A	PPB	19 Oct 2017 10:08:05	U	SMPL
ChkStd - 1	ChkStd	23569	6.9512	-	7470A	PPB	19 Oct 2017 10:10:12	U	SMPL
PB103422BL - 1	PBW	-204	-0.0655	-	7470A	PPB	19 Oct 2017 10:12:21	U	SMPL
PB103422BS - 1	LCSW	12857	3.7895	-	7470A	PPB	19 Oct 2017 10:14:30	U	SMPL
PB103422TB - 1	PB103385TB	-109	-0.0375	-	7470A	PPB	19 Oct 2017 10:16:37	U	SMPL
I5943-03 - 1	SU-03-101717	-3	-0.0062	-	7470A	PPB	19 Oct 2017 10:18:45	U	SMPL
I5944-02 - 1	CONCRETE-PILE	-5	-0.0068	-	7470A	PPB	19 Oct 2017 10:20:53	U	SMPL
I5951-03 - 1	294-COMP	7	-0.0033	-	7470A	PPB	19 Oct 2017 10:23:02	U	SMPL
I5952-03 - 1	RO-COMP	29	0.0032	-	7470A	PPB	19 Oct 2017 10:25:08	U	SMPL
CCV52 - 1	CCV52	17124	5.0489	-	7470A	PPB	19 Oct 2017 10:27:15	U	SMPL
CCB52 - 1	CCB52	-136	-0.0455	-	7470A	PPB	19 Oct 2017 10:29:22	U	SMPL
I5953-03 - 1	EO-02-101717-A	20	0.0006	-	7470A	PPB	19 Oct 2017 10:31:29	U	SMPL
I5953-06 - 1	EO-02-101717-B	17	-0.0003	-	7470A	PPB	19 Oct 2017 10:33:36	U	SMPL
I5953-06DUP - 1	EO-02-101717-BDUP	3	-0.0044	-	7470A	PPB	19 Oct 2017 10:35:43	U	SMPL
I5953-06MS - 1	EO-02-101717-BMS	12756	3.7597	-	7470A	PPB	19 Oct 2017 10:37:50	U	SMPL
I5953-06MSD - 1	EO-02-101717-BMSD	12581	3.7080	-	7470A	PPB	19 Oct 2017 10:39:58	U	SMPL
I5953-06LX5 - 1		-113	-0.0387	-	7470A	PPB	19 Oct 2017 10:42:06	U	SMPL
I5953-06A - 1		14139	4.1679	-	7470A	PPB	19 Oct 2017 10:44:14	U	SMPL
CCV53 - 1	CCV53	17105	5.0433	-	7470A	PPB	19 Oct 2017 10:46:22	U	SMPL
CCB53 - 1	CCB53	-146	-0.0484	-	7470A	PPB	19 Oct 2017 10:48:31	U	SMPL