

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: DHVANI

Date: 10/19/2017

DATE & TIME IN: 10/19/2017 11:15

OVEN TEMP: 103 °C

Run Number: LB90933

DATE & TIME OUT: 10/19/2017 15:15

OVEN TEMP: 103 °C

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB90933BL	LB90933BL	1.4227	1.4227	100	1.4227	1.4227	1.4227	0	0
2	LB90933BS	LB90933BS	1.427	1.427	100	1.4854	1.4854	1.4854	0.0584	584
3	I5960-01	COMP	1.4198	1.4198	100	1.4367	1.4367	1.4367	0.0169	169
4	I5960-01DUP	COMPDUP	1.4234	1.4234	100	1.4399	1.4399	1.4399	0.0165	165
5	I5966-03	OUTFALL001-171017	1.4388	1.4388	600	1.4389	1.4389	1.4389	0.0001	0.2
6	I5968-01	101717-TIDO-F-D-B	1.4116	1.4116	300	1.4224	1.4224	1.4224	0.0108	36
7	I5968-02	101717-TIDO-F-D-S	1.4222	1.4222	400	1.4363	1.4363	1.4363	0.0141	35.2
8	I5968-03	101717-TIDO-F-U-B	1.4339	1.4339	400	1.4543	1.4543	1.4543	0.0204	51
9	I5968-04	101717-TIDO-F-U-S	1.4259	1.4259	400	1.4407	1.4407	1.4407	0.0148	37
10	I5972-01	101617-TIDO-E-D-B	1.4255	1.4255	500	1.4344	1.4344	1.4344	0.0089	17.8
11	I5972-02	101617-TIDO-E-D-S	1.4222	1.4222	400	1.4415	1.4415	1.4415	0.0193	48.3
12	I5972-03	101617-TIDO-E-U-B	1.4203	1.4203	400	1.4331	1.4331	1.4331	0.0128	32
13	I5972-04	101617-TIDO-E-U-S	1.4176	1.4176	600	1.4297	1.4297	1.4297	0.0121	20.2

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS101917

WorkList ID : 104406

Date : 10/19/2017 9:49:08 AM

LB90933

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/25/2017	Water	I5960-01	TSS	Cool 4 deg C	ARAM01	M11	COMP	10/18/2017	SM2540 D
10/24/2017	Water	I5966-03	TSS	Cool 4 deg C	SHAW22	M41	OUTFALL001-171017	10/17/2017	SM2540 D
10/24/2017	Water	I5968-01	TSS	Cool 4 deg C	HDR103	N42	101717-TIDO-F-D-B	10/17/2017	SM2540 D
10/24/2017	Water	I5968-02	TSS	Cool 4 deg C	HDR103	N42	101717-TIDO-F-D-S	10/17/2017	SM2540 D
10/24/2017	Water	I5968-03	TSS	Cool 4 deg C	HDR103	N42	101717-TIDO-F-U-B	10/17/2017	SM2540 D
10/24/2017	Water	I5968-04	TSS	Cool 4 deg C	HDR103	N42	101717-TIDO-F-U-S	10/17/2017	SM2540 D
10/23/2017	Water	I5972-01	TSS	Cool 4 deg C	HDR103	M41	101617-TIDO-E-D-B	10/16/2017	SM2540 D
10/23/2017	Water	I5972-02	TSS	Cool 4 deg C	HDR103	M41	101617-TIDO-E-D-S	10/16/2017	SM2540 D
10/23/2017	Water	I5972-03	TSS	Cool 4 deg C	HDR103	M41	101617-TIDO-E-U-B	10/16/2017	SM2540 D
10/23/2017	Water	I5972-04	TSS	Cool 4 deg C	HDR103	M41	101617-TIDO-E-U-S	10/16/2017	SM2540 D

Date/Time 10/19/17 10:45
 Received by: DP
 Relinquished by: DP

Date/Time 10/19/17 11:25
 Received by: CP
 Relinquished by: DP