


245.1

LB 90934
INSTRUMENT ID: CV2

Linear

 μ Abs.:

31174



A= 0.0000e+000

B= 3.2266e-004

C= -4.2613e-002

Rho= 0.9998486

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	23	0.000	23				
0.2	0.200	0.173	-0.027	667	0.0 %	667				
2.5	2.500	2.561	0.061	8069	0.0 %	8069				
5.0	5.000	5.090	0.090	15906	0.0 %	15906				
7.5	7.500	7.396	-0.104	23055	0.0 %	23055				
10.0	10.000	10.016	0.016	31174	0.0 %	31174				

LB90934 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 19 Oct 2017 10:59:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	245.1	PPB	19 Oct 2017 11:01:47	S	Std
0.2 - 1	50.2	667	-	0.2000	245.1	PPB	19 Oct 2017 11:03:56	S	Std
2.5 - 1	52.5	8069	-	2.5000	245.1	PPB	19 Oct 2017 11:06:04	S	Std
5.0 - 1	55	15906	-	5.0000	245.1	PPB	19 Oct 2017 11:08:12	S	Std
7.5 - 1	57.5	23055	-	7.5000	245.1	PPB	19 Oct 2017 11:10:20	S	Std
10.0 - 1	510	31174	-	10.0000	245.1	PPB	19 Oct 2017 11:12:28	S	Std
ICV19 - 1	ICV19	12311	3.9296	-245.1		PPB	19 Oct 2017 11:15:01	U	SMPL
ICB19 - 1	ICB19	-133	-0.0855	-245.1		PPB	19 Oct 2017 11:17:07	U	SMPL
CCV54 - 1	CCV54	15878	5.0805	-245.1		PPB	19 Oct 2017 11:19:15	U	SMPL
CCB54 - 1	CCB54	-142	-0.0884	-245.1		PPB	19 Oct 2017 11:21:21	U	SMPL
CRA - 1	CRA	655	0.1687	-245.1		PPB	19 Oct 2017 11:23:29	U	SMPL
HighStd - 1	HighStd	31337	10.0685	-245.1		PPB	19 Oct 2017 11:25:36	U	SMPL
ChkStd - 1	ChkStd	22429	7.1943	-245.1		PPB	19 Oct 2017 11:27:43	U	SMPL
PB103423BL - 1	PBW	-59	-0.0617	-245.1		PPB	19 Oct 2017 11:34:36	U	SMPL
PB103423BS - 1	LCSW	12210	3.8970	-245.1		PPB	19 Oct 2017 11:36:43	U	SMPL
I5968-01 - 1	101717-TIDO-F-D-B	-111	-0.0784	-245.1		PPB	19 Oct 2017 11:38:50	U	SMPL
I5968-02 - 1	101717-TIDO-F-D-S	-2	-0.0433	-245.1		PPB	19 Oct 2017 11:40:59	U	SMPL
I5968-03 - 1	101717-TIDO-F-U-B	9	-0.0397	-245.1		PPB	19 Oct 2017 11:43:07	U	SMPL
I5968-04 - 1	101717-TIDO-F-U-S	29	-0.0333	-245.1		PPB	19 Oct 2017 11:45:15	U	SMPL
I5972-01 - 1	101617-TIDO-E-D-B	-2	-0.0433	-245.1		PPB	19 Oct 2017 11:47:22	U	SMPL
CCV55 - 1	CCV55	15734	5.0341	-245.1		PPB	19 Oct 2017 11:49:29	U	SMPL
CCB55 - 1	CCB55	-108	-0.0775	-245.1		PPB	19 Oct 2017 11:51:35	U	SMPL
I5972-02 - 1	101617-TIDO-E-D-S	-4	-0.0439	-245.1		PPB	19 Oct 2017 11:53:43	U	SMPL
I5972-03 - 1	101617-TIDO-E-U-B	-7	-0.0449	-245.1		PPB	19 Oct 2017 11:55:50	U	SMPL
I5972-04 - 1	101617-TIDO-E-U-S	-3	-0.0436	-245.1		PPB	19 Oct 2017 11:57:56	U	SMPL
I5972-04DUP - 1	101617-TIDO-E-U-SDUP	6	-0.0407	-245.1		PPB	19 Oct 2017 12:00:03	U	SMPL
I5972-04MS - 1	101617-TIDO-E-U-SMS	14352	4.5882	-245.1		PPB	19 Oct 2017 12:02:11	U	SMPL
I5972-04MSD - 1	101617-TIDO-E-U-SMSD	14966	4.7863	-245.1		PPB	19 Oct 2017 12:04:18	U	SMPL
I5972-04LX5 - 1		-122	-0.0820	-245.1		PPB	19 Oct 2017 12:06:26	U	SMPL
I5972-04A - 1		15700	5.0231	-245.1		PPB	19 Oct 2017 12:08:34	U	SMPL
CCV56 - 1	CCV56	15704	5.0244	-245.1		PPB	19 Oct 2017 12:10:43	U	SMPL
CCB56 - 1	CCB56	-139	-0.0875	-245.1		PPB	19 Oct 2017 12:12:50	U	SMPL