



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : _____

Run Number: LB90935

Temp.: 20.3 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/19/2017	11:50
2	CAL2	1	Water			7.00	10/19/2017	11:51
3	CAL3	1	Water			10.03	10/19/2017	11:52
4	ICV	1	Water			6.99	10/19/2017	12:00
5	CCV1	1	Water			2.01	10/19/2017	12:01
6	I5994-01	1	Water			4.27	10/19/2017	12:03
7	I5994-01DUP	1	Water			4.28	10/19/2017	12:05
8	CCV2	1	Water			12.01	10/19/2017	12:06

1390933

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5994

WorkList ID : 104414

Date : 10/19/2017 10:39:18 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/16/2017	Water	I5994-01-17	pH	Cool 4 deg C	STER11	M11	POOLWATER	10/16/2017	SM 4500-PH B

Date/Time 10-19-17 10:40 AM
Received by: JP
Relinquished by: JP

Date/Time 10-19-17 11:55 AM
Received by: JP
Relinquished by: JP