


7471A

LB 90936
INSTRUMENT ID: CV4

Linear

 μ Abs.:

4177E



A= 0.0000e+000

B= 2.4107e-004

C= -7.5705e-003

Rho= 0.9999168

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	39	0.000	39				
0.2	0.200	0.216	0.016	928	0.0 %	928				
2.5	2.500	2.541	0.041	10573	0.0 %	10573				
5.0	5.000	4.934	-0.066	20497	0.0 %	20497				
7.5	7.500	7.444	-0.056	30908	0.0 %	30908	0			
10.0	10.000	10.064	0.064	41776	0.0 %	41776				

LB90936 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 19 Oct 2017 12:21:16

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.00007471A	PPB		19 Oct 2017 12:24:43	S	Std
0.2 - 1	50.2	928	-	0.20007471A	PPB		19 Oct 2017 12:27:01	S	Std
2.5 - 1	52.5	10573	-	2.50007471A	PPB		19 Oct 2017 12:29:19	S	Std
5.0 - 1	55	20497	-	5.00007471A	PPB		19 Oct 2017 12:31:36	S	Std
7.5 - 1	57.5	30908	-	7.50007471A	PPB		19 Oct 2017 12:33:53	S	Std
10.0 - 1	510	41776	-	10.00007471A	PPB		19 Oct 2017 12:38:46	S	Std
ICV20 - 1	ICV20	16474	3.9639	-7471A	PPB		19 Oct 2017 12:42:58	U	SMPL
ICB20 - 1	ICB20	-158	-0.0457	-7471A	PPB		19 Oct 2017 12:45:14	U	SMPL
CCV57 - 1	CCV57	20284	4.8824	-7471A	PPB		19 Oct 2017 12:47:30	U	SMPL
CCB57 - 1	CCB57	-185	-0.0522	-7471A	PPB		19 Oct 2017 12:49:46	U	SMPL
CRA - 1	CRA	859	0.1995	-7471A	PPB		19 Oct 2017 12:52:02	U	SMPL
HighStd - 1	HighStd	41738	10.0543	-7471A	PPB		19 Oct 2017 12:54:18	U	SMPL
ChkStd - 1	ChkStd	30106	7.2502	-7471A	PPB		19 Oct 2017 12:56:34	U	SMPL
PB103426BL - 1	PBS	-275	-0.0739	-7471A	PPB		19 Oct 2017 12:58:53	U	SMPL
PB103426BS - 1	LCSS	14758	3.5502	-7471A	PPB		19 Oct 2017 13:01:11	U	SMPL
I5950-01 - 1	20170718-C-COMP1	98984	23.8548	-7471A	PPB		19 Oct 2017 13:03:28	U	SMPL
I5950-02 - 1	20170719-C-COMP2	99025	23.8647	-7471A	PPB		19 Oct 2017 13:05:45	U	SMPL
I5950-03 - 1	20170720-C-COMP3	98043	23.6280	-7471A	PPB		19 Oct 2017 13:08:05	U	SMPL
I5950-04 - 1	20170721-PA-COMP1	49792	11.9959	-7471A	PPB		19 Oct 2017 13:10:28	U	SMPL
I5950-05 - 1	20170722-C-A-COMP1	82125	19.7906	-7471A	PPB		19 Oct 2017 13:12:47	U	SMPL
CCV58 - 1	CCV58	19549	4.7052	-7471A	PPB		19 Oct 2017 13:15:05	U	SMPL
CCB58 - 1	CCB58	-213	-0.0589	-7471A	PPB		19 Oct 2017 13:17:24	U	SMPL
I5950-06 - 1	20170723-C-A-COMP2	76971	18.5481	-7471A	PPB		19 Oct 2017 13:19:41	U	SMPL
I5950-07 - 1	20170724-C-A-COMP3	100314	24.1754	-7471A	PPB		19 Oct 2017 13:21:57	U	SMPL
I5950-08 - 1	20170725-PA-A-COMP1	44970	10.8335	-7471A	PPB		19 Oct 2017 13:24:15	U	SMPL
I5981-01 - 1	NB-01-101817-A	3441	0.8220	-7471A	PPB		19 Oct 2017 13:26:35	U	SMPL
I5981-04 - 1	NB-01-101817-B	1277	0.3003	-7471A	PPB		19 Oct 2017 13:28:54	U	SMPL
I5981-07 - 1	NB-01-101817-C	988	0.2306	-7471A	PPB		19 Oct 2017 13:31:11	U	SMPL
I5984-01 - 1	20-02	25781	6.2075	-7471A	PPB		19 Oct 2017 13:33:28	U	SMPL
I5988-01 - 1	RBR251373-78	958903	HIGH	-7471A	PPB		19 Oct 2017 13:35:45	X	
I5988-01DUP - 1	RBR251373-78DUP	915363	HIGH	-7471A	PPB		19 Oct 2017 13:37:38	X	
I5988-01MS - 1	RBR251373-78MS	946812	HIGH	-7471A	PPB		19 Oct 2017 13:40:39	X	
CCV59 - 1	CCV59	21560	5.1900	-7471A	PPB		19 Oct 2017 13:42:26	U	SMPL
CCB59 - 1	CCB59	-279	-0.0748	-7471A	PPB		19 Oct 2017 13:44:44	U	SMPL
I5988-01MSD - 1	RBR251373-78MSD	955053	HIGH	-7471A	PPB		19 Oct 2017 13:45:45	X	
I5988-01LX5 - 1		352483	HIGH	-7471A	PPB		19 Oct 2017 13:56:37	X	
I5988-01A - 1		943817	HIGH	-7471A	PPB		19 Oct 2017 13:59:40	X	
I5950-01DLX5 - 1		18414	4.4316	-7471A	PPB		19 Oct 2017 14:01:50	U	SMPL
I5950-02DLX5 - 1		20148	4.8496	-7471A	PPB		19 Oct 2017 14:05:08	U	SMPL
I5950-03DLX5 - 1		20402	4.9108	-7471A	PPB		19 Oct 2017 14:07:27	U	SMPL
I5950-04DLX2 - 1		25509	6.1420	-7471A	PPB		19 Oct 2017 14:09:46	U	SMPL
I5950-05DLX5 - 1		16502	3.9706	-7471A	PPB		19 Oct 2017 14:12:05	U	SMPL
I5950-06DLX5 - 1		15562	3.7440	-7471A	PPB		19 Oct 2017 14:14:25	U	SMPL
I5950-07DLX5 - 1		20657	4.9723	-7471A	PPB		19 Oct 2017 14:16:41	U	SMPL
CCV60 - 1	CCV60	20763	4.9978	-7471A	PPB		19 Oct 2017 14:18:58	U	SMPL
CCB60 - 1	CCB60	-201	-0.0560	-7471A	PPB		19 Oct 2017 14:21:15	U	SMPL
I5950-08DLX2 - 1		23343	5.6198	-7471A	PPB		19 Oct 2017 14:23:32	U	SMPL
I5988-01DLX10 - 1		241563	58.2268	-7471A	PPB		19 Oct 2017 14:37:00	U	SMPL
I5988-01DUPDLX10 - 1		227415	54.8161	-7471A	PPB		19 Oct 2017 14:39:16	U	SMPL
I5988-01MSDLX10 - 1		248117	59.8068	-7471A	PPB		19 Oct 2017 14:42:00	U	SMPL
I5988-01MSDDLX10 - 1		247450	59.6460	-7471A	PPB		19 Oct 2017 14:44:47	U	SMPL
I5988-01LDLX50 - 1		49141	11.8390	-7471A	PPB		19 Oct 2017 14:47:37	U	SMPL
I5988-01ADLX10 - 1		240481	57.9659	-7471A	PPB		19 Oct 2017 14:50:25	U	SMPL
CCV61 - 1	CCV61	21255	5.1164	-7471A	PPB		19 Oct 2017 14:52:45	U	SMPL
CCB61 - 1	CCB61	-240	-0.0654	-7471A	PPB		19 Oct 2017 14:55:27	U	SMPL
I5988-01DL2X100 - 1		25987	6.2572	-7471A	PPB		19 Oct 2017 14:57:45	U	SMPL
I5988-01DUPDL2X100 - 1		24419	5.8792	-7471A	PPB		19 Oct 2017 15:00:02	U	SMPL
I5988-01MSDL2X100 - 1		27431	6.6053	-7471A	PPB		19 Oct 2017 15:02:19	U	SMPL
I5988-01MSDDL2X100 - 1		27230	6.5569	-7471A	PPB		19 Oct 2017 15:04:37	U	SMPL
I5988-01LDL2X500 - 1		4832	1.1573	-7471A	PPB		19 Oct 2017 15:06:56	U	SMPL
I5988-01ADL2X100 - 1		26104	6.2854	-7471A	PPB		19 Oct 2017 15:09:14	U	SMPL
CCV62 - 1	CCV62	22016	5.2999	-7471A	PPB		19 Oct 2017 15:11:31	U	SMPL
CCB62 - 1	CCB62	4	-0.0066	-7471A	PPB		19 Oct 2017 15:13:49	U	SMPL