



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90956

Temp.: 22.5 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/19/2017	17:30
2	CAL2	1	Water			7.00	10/19/2017	17:31
3	CAL3	1	Water			10.02	10/19/2017	17:32
4	ICV	1	Water			7.00	10/19/2017	17:33
5	CCV1	1	Water			2.01	10/19/2017	17:35
6	I6010-01	1	Solid	20.02	20	6.70	10/19/2017	17:36
7	I6010-01DUP	1	Solid	20.03	20	6.71	10/19/2017	17:37
8	I6013-01	1	Solid	20.05	20	6.40	10/19/2017	17:38
9	CCV2	1	Water			12.02	10/19/2017	17:41

15790956

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6010

WorkList ID : 104426

Date : 10/19/2017 5:30:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/18/2017	Solid	I6010-01 -17	pH	Cool 4 deg C	TULL02	N41	FK-PS-01-018	10/18/2017	9045C
10/19/2017	Solid	I6013-01 -17	pH	Cool 4 deg C	TULL02	N41	FK-PS-01-019	10/19/2017	9045C

Date/Time 10-19-17 5:30 PM
Received by: JD
Relinquished by: JD

Date/Time 10-19-17 6:35 PM
Received by: CD
Relinquished by: JD