



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90958

Temp.: 22.2 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/19/2017	17:10
2	CAL2	1	Water			7.01	10/19/2017	17:11
3	CAL3	1	Water			10.03	10/19/2017	17:13
4	ICV	1	Water			6.99	10/19/2017	17:15
5	CCV1	1	Water			2.01	10/19/2017	17:16
6	I6017-04	1	Solid	20.02	20	7.54	10/19/2017	17:19
7	I6017-04DUP	1	Solid	20.03	20	7.55	10/19/2017	17:20
8	I6017-08	1	Solid	20.01	20	7.28	10/19/2017	17:21
9	I6018-03	1	Solid	20.03	20	8.86	10/19/2017	17:22
10	CCV2	1	Water			12.01	10/19/2017	17:25

WORKLIST(Hardcopy Internal Chain)

7890950-2B90958
JD

WorkList Name : CORROSIONITY I6018 WorkList ID : 104425

Date : 10/19/2017 5:15:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2017	Solid	I6017-04	Corrosivity	Cool 4 deg C	PSEG01	M52	COMP-1	10/19/2017	9045C
10/24/2017	Solid	I6017-08	Corrosivity	Cool 4 deg C	PSEG01	M52	RT3867-RT5158-COMP-2	10/19/2017	9045C
10/24/2017	Solid	I6018-03	Corrosivity	Cool 4 deg C	PSEG05	M52	TR-05-101917	10/19/2017	9045C

Date/Time 10/19/17 4:00pm
Received by: JH
Relinquished by: [Signature]

Date/Time 10-19-17 5:10pm
Received by: [Signature]
Relinquished by: JH