


7470A

LB 90961
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34999



A= 0.0000e+000

B= 2.8804e-004

C= 2.7904e-002

Rho= 0.9997914

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.032	0.032	15	0.000	15				
0.2	0.200	0.235	0.035	718	0.0 %	718				
2.5	2.500	2.519	0.019	8649	0.0 %	8649				
5.0	5.000	4.886	-0.114	16867	0.0 %	16867				
7.5	7.500	7.418	-0.082	25658	0.0 %	25658				
10.0	10.000	10.109	0.109	34999	0.0 %	34999				

LB90961 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 20 Oct 2017 09:35:50				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	15	-	0.0000	7470A	PPB	20 Oct 2017 10:23:41	S Std
0.2 - 1	50.2	718	-	0.2000	7470A	PPB	20 Oct 2017 10:25:49	S Std
2.5 - 1	52.5	8649	-	2.5000	7470A	PPB	20 Oct 2017 10:27:58	S Std
5.0 - 1	55	16867	-	5.0000	7470A	PPB	20 Oct 2017 10:30:06	S Std
7.5 - 1	57.5	25658	-	7.5000	7470A	PPB	20 Oct 2017 10:32:14	S Std
10.0 - 1	510	34999	-	10.0000	7470A	PPB	20 Oct 2017 10:34:22	S Std
ICV21 - 1	ICV21	12929	3.7520	-	7470A	PPB	20 Oct 2017 10:37:04	U SMPL
ICB21 - 1	ICB21	-164	-0.0193	-	7470A	PPB	20 Oct 2017 10:39:10	U SMPL
CCV65 - 1	CCV65	16971	4.9163	-	7470A	PPB	20 Oct 2017 10:41:17	U SMPL
CCB65 - 1	CCB65	-168	-0.0205	-	7470A	PPB	20 Oct 2017 10:43:24	U SMPL
CRA - 1	CRA	683	0.2246	-	7470A	PPB	20 Oct 2017 10:45:32	U SMPL
HighStd - 1	HighStd	35229	10.1753	-	7470A	PPB	20 Oct 2017 10:47:38	U SMPL
ChkStd - 1	ChkStd	24073	6.9619	-	7470A	PPB	20 Oct 2017 10:49:45	U SMPL
PB103462BL - 1	PBW	-244	-0.0424	-	7470A	PPB	20 Oct 2017 10:51:55	U SMPL
PB103462BS - 1	LCSW	13506	3.9182	-	7470A	PPB	20 Oct 2017 10:54:04	U SMPL
PB103462TB - 1	PB103424TB	-118	-0.0061	-	7470A	PPB	20 Oct 2017 10:56:11	U SMPL
I5981-03 - 1	NB-01-101817-A	-7	0.0259	-	7470A	PPB	20 Oct 2017 10:58:20	U SMPL
I5981-06 - 1	NB-01-101817-B	0	0.0279	-	7470A	PPB	20 Oct 2017 11:00:28	U SMPL
I5981-09 - 1	NB-01-101817-C	0	0.0279	-	7470A	PPB	20 Oct 2017 11:02:37	U SMPL
I5984-02 - 1	20-02	-8	0.0256	-	7470A	PPB	20 Oct 2017 11:04:44	U SMPL
CCV66 - 1	CCV66	17056	4.9407	-	7470A	PPB	20 Oct 2017 11:06:50	U SMPL
CCB66 - 1	CCB66	-183	-0.0248	-	7470A	PPB	20 Oct 2017 11:08:57	U SMPL
I5988-03 - 1	RBR251373-78	297	0.1135	-	7470A	PPB	20 Oct 2017 11:11:05	U SMPL
I5988-03DUP - 1	RBR251373-78DUP	351	0.1290	-	7470A	PPB	20 Oct 2017 11:13:11	U SMPL
I5988-03MS - 1	RBR251373-78MS	13050	3.7869	-	7470A	PPB	20 Oct 2017 11:15:18	U SMPL
I5988-03MSD - 1	RBR251373-78MSD	12941	3.7555	-	7470A	PPB	20 Oct 2017 11:17:25	U SMPL
PB103463BL - 1	PBW	-142	-0.0130	-	7470A	PPB	20 Oct 2017 11:19:32	U SMPL
PB103463BS - 1	LCSW	12413	3.6034	-	7470A	PPB	20 Oct 2017 11:21:41	U SMPL
PB103463TB - 1	PB103409TB	-132	-0.0101	-	7470A	PPB	20 Oct 2017 11:23:49	U SMPL
I5932-04 - 1	W-2	-40	0.0164	-	7470A	PPB	20 Oct 2017 11:25:57	U SMPL
I5950-09 - 1	20170722-C-A-COMP1	4999	1.4678	-	7470A	PPB	20 Oct 2017 11:28:05	U SMPL
I5950-10 - 1	20170723-C-A-COMP2	8669	2.5249	-	7470A	PPB	20 Oct 2017 11:30:12	U SMPL
CCV67 - 1	CCV67	17023	4.9312	-	7470A	PPB	20 Oct 2017 11:32:19	U SMPL
CCB67 - 1	CCB67	-192	-0.0274	-	7470A	PPB	20 Oct 2017 11:34:26	U SMPL
I5950-11 - 1	20170724-C-A-COMP3	10589	3.0780	-	7470A	PPB	20 Oct 2017 11:36:34	U SMPL
I5950-12 - 1	20170725-PA-A-COMP1	3209	0.9522	-	7470A	PPB	20 Oct 2017 11:38:41	U SMPL
I5950-12DUP - 1	20170725-PA-A-COMP1DUP	3274	0.9710	-	7470A	PPB	20 Oct 2017 11:40:48	U SMPL
I5950-12MS - 1	20170725-PA-A-COMP1MS	17999	5.2124	-	7470A	PPB	20 Oct 2017 11:42:55	U SMPL
I5950-12MSD - 1	20170725-PA-A-COMP1MSD	17728	5.1343	-	7470A	PPB	20 Oct 2017 11:45:02	U SMPL
I5988-03LX5 - 1		-116	-0.0055	-	7470A	PPB	20 Oct 2017 11:47:11	U SMPL
I5988-03A - 1		13820	4.0086	-	7470A	PPB	20 Oct 2017 11:49:20	U SMPL
I5950-12LX5 - 1		468	0.1627	-	7470A	PPB	20 Oct 2017 11:51:28	U SMPL
I5950-12A - 1		18013	5.2164	-	7470A	PPB	20 Oct 2017 11:53:37	U SMPL
CCV68 - 1	CCV68	16768	4.8578	-	7470A	PPB	20 Oct 2017 11:55:44	U SMPL
CCB68 - 1	CCB68	-180	-0.0239	-	7470A	PPB	20 Oct 2017 11:57:52	U SMPL