

ISM02.4


LB 90 972

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41058



A= 0.0000e+000

B= 2.4236e-004

C= -2.0571e-002

Rho= 0.9999187

Accept=Accepted

slope
y=mx+intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.00	0.000	-0.015	-0.015	25	0.000	25					-
0.2	0.200	0.192	-0.008	877	0.0 %	877					-4
2.5	2.500	2.500	0.000	10401	0.0 %	10401					0
5.0	5.000	5.002	0.002	20724	0.0 %	20724					0
7.5	7.500	7.590	0.090	31403	0.0 %	31403					1
10.0	10.000	9.930	-0.070	41058	0.0 %	41058					-1

NB
10/20/18

LB90972
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	25	-		0 ISM02.4	PPB	10/20/2017 14:51	Std
	0.2 S0.2	877	-		0.2 ISM02.4	PPB	10/20/2017 14:54	Std
	2.5 S2.5	10401	-		2.5 ISM02.4	PPB	10/20/2017 14:56	Std
	5 S5	20724	-		5 ISM02.4	PPB	10/20/2017 14:58	Std
	7.5 S7.5	31403	-		7.5 ISM02.4	PPB	10/20/2017 15:00	Std
	10 S10	41058	-		10 ISM02.4	PPB	10/20/2017 15:03	Std
ICV	ICV	15418	3.7161 -		ISM02.4	PPB	10/20/2017 15:06	SMPL
ICB	ICB	-136	-0.0535 -		ISM02.4	PPB	10/20/2017 15:08	SMPL
CCV069	CCV069	20787	5.0173 -		ISM02.4	PPB	10/20/2017 15:10	SMPL
CCB069	CCB069	-152	-0.0574 -		ISM02.4	PPB	10/20/2017 15:12	SMPL
PB103445BL	PBW003	0	-0.0206 -		ISM02.4	PPB	10/20/2017 15:15	SMPL
I5958-12	TWP-B-67	1715	0.3951 -		ISM02.4	PPB	10/20/2017 15:17	SMPL
I5976-03	EQB-4-101817	77	-0.0019 -		ISM02.4	PPB	10/20/2017 15:19	SMPL
I5978-01	TWP-B-7	2002	0.4646 -		ISM02.4	PPB	10/20/2017 15:21	SMPL
I5978-02	TWP-B-9	24022	5.8013 -		ISM02.4	PPB	10/20/2017 15:24	SMPL
CCV070	CCV070	20730	5.0035 -		ISM02.4	PPB	10/20/2017 15:26	SMPL
CCB070	CCB070	-184	-0.0652 -		ISM02.4	PPB	10/20/2017 15:28	SMPL