

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB90983

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP58934
calibration std. phosphate 0.5 ppm	WP58933
calibration std. phosphate 0.3 ppm	WP58932
calibration std. 0 ppm	WP58929
phosphate ICV std.	WP58936
phosphate CCV std.	WP58935
calibration std. phosphate 0.1 ppm	WP58931
calibration std. phosphate 0.05 ppm	WP58930
Combined reagent	WP58939
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036

Intercept: -0.0008Slope: 0.6506Regression: 0.99981

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	10/20/2017	13:45
2	CAL2	0.05	1	50	50	0.033	10/20/2017	13:45
3	CAL3	0.10	1	50	50	0.065	10/20/2017	13:46
4	CAL4	0.30	1	50	50	0.186	10/20/2017	13:46
5	CAL5	0.50	1	50	50	0.331	10/20/2017	13:47
6	CAL6	1.00	1	50	50	0.649	10/20/2017	13:47



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB90983

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.501	10/20/2017	13:48
2	ICB		1	50	50	0.000	0.001	10/20/2017	13:48
3	CCV1	0.50	1	50	50	0.328	0.505	10/20/2017	13:49
4	CCB1		1	50	50	0.000	0.001	10/20/2017	13:49
5	PB103498BL		1	50	50	0.000	0.001	10/20/2017	13:50
6	PB103498BS	0.50	1	50	50	0.327	0.504	10/20/2017	13:50
7	I5999-01		1	50	50	1.336	2.055	10/20/2017	13:51
8	I5999-01DUP		1	50	50	1.325	2.038	10/20/2017	13:51
9	I5999-01MS	0.50	1	50	50	1.412	2.172	10/20/2017	13:52
10	I5999-01MSD	0.50	1	50	50	1.410	2.168	10/20/2017	13:52
11	I5999-01		10	50	50	0.321	0.495	10/20/2017	14:00
12	I5999-01DUP		10	50	50	0.318	0.490	10/20/2017	14:01
13	I5999-01MS	0.50	10	50	50	0.348	0.536	10/20/2017	14:01
14	I5999-01MSD	0.50	10	50	50	0.350	0.539	10/20/2017	14:02
15	CCV2	0.50	1	50	50	0.327	0.504	10/20/2017	14:02
16	CCB2		1	50	50	0.000	0.001	10/20/2017	14:03