

7471A


LB90987

INSTRUMENT ID: CV8

Linear

 μ Abs.:

40314



A= 0.0000e+000

B= 2.5084e-004

C= -1.5464e-002

Rho= 0.9997893

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	47	0.000	47				
0.2	0.200	0.222	0.022	946	0.0 %	946				
2.5	2.500	2.577	0.077	10337	0.0 %	10337				
5.0	5.000	4.903	-0.097	19608	0.0 %	19608				
7.5	7.500	7.404	-0.096	29580	0.0 %	29580	0			
10.0	10.000	10.097	0.097	40314	0.0 %	40314				

LB90987 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 20 Oct 2017 18:14:07				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	30	47	-	0.00007471A	PPB	20 Oct 2017 18:16:27	S	Std
0.2 - 1	50.2	946	-	0.20007471A	PPB	20 Oct 2017 18:18:45	S	Std
2.5 - 1	52.5	10337	-	2.50007471A	PPB	20 Oct 2017 18:21:02	S	Std
5.0 - 1	55	19608	-	5.00007471A	PPB	20 Oct 2017 18:23:20	S	Std
7.5 - 1	57.5	29580	-	7.50007471A	PPB	20 Oct 2017 18:25:37	S	Std
10.0 - 1	510	40314	-	10.00007471A	PPB	20 Oct 2017 18:30:30	S	Std
ICV22 - 1	ICV22	15106	3.7737	-7471A	PPB	20 Oct 2017 18:33:08	U	SMPL
ICB22 - 1	ICB22	-165	-0.0569	-7471A	PPB	20 Oct 2017 18:35:25	U	SMPL
CCV71 - 1	CCV71	19614	4.9045	-7471A	PPB	20 Oct 2017 18:37:41	U	SMPL
CCB71 - 1	CCB71	-167	-0.0574	-7471A	PPB	20 Oct 2017 18:39:57	U	SMPL
CRA - 1	CRA	935	0.2191	-7471A	PPB	20 Oct 2017 18:42:14	U	SMPL
HighStd - 1	HighStd	39907	9.9948	-7471A	PPB	20 Oct 2017 18:44:30	U	SMPL
ChkStd - 1	ChkStd	26866	6.7236	-7471A	PPB	20 Oct 2017 18:46:45	U	SMPL
PB103490BL - 1	PBS	-206	-0.0671	-7471A	PPB	20 Oct 2017 18:49:03	U	SMPL
PB103490BS - 1	LCSS	15405	3.8487	-7471A	PPB	20 Oct 2017 18:53:56	U	SMPL
I6008-01 - 1	LT-TS-SPLP-1	3454	0.8509	-7471A	PPB	20 Oct 2017 18:56:12	U	SMPL
I6008-02 - 1	LT-TS-SPLP-2	3490	0.8600	-7471A	PPB	20 Oct 2017 18:58:30	U	SMPL
I6008-03 - 1	LT-TS-SPLP-3	11300	2.8190	-7471A	PPB	20 Oct 2017 19:00:47	U	SMPL
I6010-01 - 1	FK-PS-01-018	456	0.0989	-7471A	PPB	20 Oct 2017 19:03:05	U	SMPL
CCV72 - 1	CCV72	19655	4.9148	-7471A	PPB	20 Oct 2017 19:07:37	U	SMPL
CCB72 - 1	CCB72	-167	-0.0574	-7471A	PPB	20 Oct 2017 19:09:53	U	SMPL
I6010-01MS - 1	FK-PS-01-018MS	15809	3.9501	-7471A	PPB	20 Oct 2017 19:12:10	U	SMPL
I6010-01MSD - 1	FK-PS-01-018MSD	15995	3.9967	-7471A	PPB	20 Oct 2017 19:14:26	U	SMPL
I6013-01 - 1	FK-PS-01-019	313	0.0630	-7471A	PPB	20 Oct 2017 19:16:42	U	SMPL
I6017-01 - 1	COMP-1	2170	0.5289	-7471A	PPB	20 Oct 2017 19:18:58	U	SMPL
I6017-05 - 1	RT3867-RT5158-COMP-2	1564	0.3769	-7471A	PPB	20 Oct 2017 19:21:15	U	SMPL
I6018-01 - 1	TR-05-101917	1443	0.3465	-7471A	PPB	20 Oct 2017 19:23:32	U	SMPL
I6010-01DUP - 1	FK-PS-01-018DUP	362	0.0753	-7471A	PPB	20 Oct 2017 19:28:36	U	SMPL
I6010-01LX5 - 1		99	0.0094	-7471A	PPB	20 Oct 2017 19:30:53	U	SMPL
I6010-01A - 1		16318	4.0778	-7471A	PPB	20 Oct 2017 19:33:11	U	SMPL
CCV73 - 1	CCV73	19655	4.9148	-7471A	PPB	20 Oct 2017 19:35:27	U	SMPL
CCB73 - 1	CCB73	-168	-0.0576	-7471A	PPB	20 Oct 2017 19:37:44	U	SMPL