



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90994

Temp.: 21.0 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
ACETIC ACID, GLACIAL, ACS, 2.5L	W2138
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/20/2017	17:00
2	CAL2	1	Water			7.01	10/20/2017	17:01
3	CAL3	1	Water			10.03	10/20/2017	17:02
4	ICV	1	Water			6.99	10/20/2017	17:03
5	CCV1	1	Water			2.01	10/20/2017	17:05
6	I6034-10	1	Water			6.54	10/20/2017	17:06
7	I6034-10DUP	1	Water			6.55	10/20/2017	17:08
8	CCV2	1	Water			12.01	10/20/2017	17:10

WORKLIST(Hardcopy Internal Chain)

2B 90994

WorkList Name : CORROSIONITY I5934

WorkList ID : 104497

Date : 10/20/2017 4:10:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2017	Water	I6034-10	Corrosivity	Cool 4 deg C	PSEG01	M52	BAKER-TANK	10/19/2017	9040C

Date/Time 10/20/17 4:10 PM
 Received by: JP
 Relinquished by: JP

Date/Time 10/20/17 5:30 PM
 Received by: CP
 Relinquished by: JP