


7470A

LB 91001
INSTRUMENT ID: CV2

Linear

 μ Abs.:

34929



A= 0.0000e+000

B= 2.8497e-004

C= -1.0270e-002

Rho= 0.9998472

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	61	0.000	61				
0.2	0.200	0.211	0.011	777	0.0 %	777				
2.5	2.500	2.487	-0.013	8763	0.0 %	8763				
5.0	5.000	4.926	-0.074	17321	0.0 %	17321				
7.5	7.500	7.626	0.126	26797	0.0 %	26797				
10.0	10.000	9.943	-0.057	34929	0.0 %	34929				

LB91001 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Oct 2017 09:44:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	61	-	0.0000	7470A	PPB	23 Oct 2017 09:53:29	S	Std
0.2 - 1	50.2	777	-	0.2000	7470A	PPB	23 Oct 2017 09:55:38	S	Std
2.5 - 1	52.5	8763	-	2.5000	7470A	PPB	23 Oct 2017 09:57:46	S	Std
5.0 - 1	55	17321	-	5.0000	7470A	PPB	23 Oct 2017 09:59:54	S	Std
7.5 - 1	57.5	26797	-	7.5000	7470A	PPB	23 Oct 2017 10:02:02	S	Std
10.0 - 1	510	34929	-	10.0000	7470A	PPB	23 Oct 2017 10:04:11	S	Std
ICV23 - 1	ICV23	13902	3.9513	-	7470A	PPB	23 Oct 2017 10:07:00	U	SMPL
ICB23 - 1	ICB23	-193	-0.0653	-	7470A	PPB	23 Oct 2017 10:09:06	U	SMPL
CCV74 - 1	CCV74	18381	5.2277	-	7470A	PPB	23 Oct 2017 10:11:14	U	SMPL
CCB74 - 1	CCB74	-187	-0.0636	-	7470A	PPB	23 Oct 2017 10:13:20	U	SMPL
CRA - 1	CRA	805	0.2191	-	7470A	PPB	23 Oct 2017 10:15:28	U	SMPL
HighStd - 1	HighStd	35490	10.1032	-	7470A	PPB	23 Oct 2017 10:17:35	U	SMPL
ChkStd - 1	ChkStd	24029	6.8372	-	7470A	PPB	23 Oct 2017 10:19:41	U	SMPL
PB103503BL - 1	PBW	-258	-0.0838	-	7470A	PPB	23 Oct 2017 10:21:50	U	SMPL
PB103503BS - 1	LCSW	12959	3.6826	-	7470A	PPB	23 Oct 2017 10:23:59	U	SMPL
PB103503TB - 1	PB103465TB	-133	-0.0482	-	7470A	PPB	23 Oct 2017 10:26:07	U	SMPL
I6017-04 - 1	COMP-1	3	-0.0094	-	7470A	PPB	23 Oct 2017 10:28:15	U	SMPL
I6017-08 - 1	RT3867-RT5158-COMP-2	19	-0.0049	-	7470A	PPB	23 Oct 2017 10:30:24	U	SMPL
I6018-03 - 1	TR-05-101917	1	-0.0100	-	7470A	PPB	23 Oct 2017 10:32:32	U	SMPL
I6018-03DUP - 1	TR-05-101917DUP	8	-0.0080	-	7470A	PPB	23 Oct 2017 10:34:39	U	SMPL
CCV75 - 1	CCV75	17386	4.9441	-	7470A	PPB	23 Oct 2017 10:36:46	U	SMPL
CCB75 - 1	CCB75	-194	-0.0656	-	7470A	PPB	23 Oct 2017 10:38:52	U	SMPL
I6018-03MS - 1	TR-05-101917MS	13163	3.7407	-	7470A	PPB	23 Oct 2017 10:41:00	U	SMPL
I6018-03MSD - 1	TR-05-101917MSD	13156	3.7387	-	7470A	PPB	23 Oct 2017 10:43:06	U	SMPL
PB103504BL - 1	PBW	-171	-0.0590	-	7470A	PPB	23 Oct 2017 10:45:13	U	SMPL
PB103504BS - 1	LCSW	13551	3.8513	-	7470A	PPB	23 Oct 2017 10:47:21	U	SMPL
I5993-01 - 1	WC-101917-LIQUID	83	0.0134	-	7470A	PPB	23 Oct 2017 10:49:29	U	SMPL
I5993-01DUP - 1	WC-101917-LIQUIDDUP	-3	-0.0111	-	7470A	PPB	23 Oct 2017 10:51:36	U	SMPL
I5993-01MS - 1	WC-101917-LIQUIDMS	15709	4.4663	-	7470A	PPB	23 Oct 2017 10:53:44	U	SMPL
I5993-01MSD - 1	WC-101917-LIQUIDMSD	15380	4.3725	-	7470A	PPB	23 Oct 2017 10:55:52	U	SMPL
I6019-12 - 1	MW-01	4735	1.3390	-	7470A	PPB	23 Oct 2017 10:58:01	U	SMPL
I6025-07 - 1	MW-01	7588	2.1520	-	7470A	PPB	23 Oct 2017 11:00:08	U	SMPL
CCV76 - 1	CCV76	17793	5.0601	-	7470A	PPB	23 Oct 2017 11:02:15	U	SMPL
CCB76 - 1	CCB76	-220	-0.0730	-	7470A	PPB	23 Oct 2017 11:04:22	U	SMPL
I6025-08 - 1	MW-02	2373	0.6660	-	7470A	PPB	23 Oct 2017 11:06:30	U	SMPL
I6025-09 - 1	MW-03	1033	0.2841	-	7470A	PPB	23 Oct 2017 11:08:36	U	SMPL
I6025-10 - 1	MW-01	38	0.0006	-	7470A	PPB	23 Oct 2017 11:10:43	U	SMPL
I6025-11 - 1	MW-02	-9	-0.0128	-	7470A	PPB	23 Oct 2017 11:12:50	U	SMPL
I6025-12 - 1	MW-03	36	-0.0000	-	7470A	PPB	23 Oct 2017 11:14:58	U	SMPL
I6034-09 - 1	BAKER-TANK	153	0.0333	-	7470A	PPB	23 Oct 2017 11:17:05	U	SMPL
I6018-03LX5 - 1		55	0.0054	-	7470A	PPB	23 Oct 2017 11:19:13	U	SMPL
I6018-03A - 1		13792	3.9200	-	7470A	PPB	23 Oct 2017 11:21:21	U	SMPL
I5993-01LX5 - 1		-48	-0.0239	-	7470A	PPB	23 Oct 2017 11:23:30	U	SMPL
I5993-01A - 1		16424	4.6700	-	7470A	PPB	23 Oct 2017 11:25:38	U	SMPL
CCV77 - 1	CCV77	17478	4.9704	-	7470A	PPB	23 Oct 2017 11:27:45	U	SMPL
CCB77 - 1	CCB77	-175	-0.0601	-	7470A	PPB	23 Oct 2017 11:29:52	U	SMPL