



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB91006
Analysis Date: 10/23/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 10/23/2017
Extraction IN Time: 10:20
Extraction OUT Time: 11:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB91006BL	LB91006BL	WATER	2	1000	100	3.1697	3.1697	3.03	3.1697	3.1697	0	0
2	LB91006BS	LB91006BS	WATER	2	1000	100	3.0225	3.0225	3.04	3.0414	3.0414	0.0189	18.9
3	LB91006BSD	LB91006BSD	WATER	2	1000	100	3.0847	3.0847	3.01	3.1037	3.1037	0.019	19
4	I5960-02	GRAB	WATER	2	1000	100	3.0491	3.0491	3.02	3.0557	3.0557	0.0066	6.6
5	I6020-01	GRAB	WATER	2	1000	100	2.9557	2.9557	3.05	2.9961	2.9961	0.0404	40.4
6	I6034-09	BAKER-TANK	WATER	2	1000	100	2.9463	2.9463	3.04	3.6587	3.6587	0.7124	712.4
7	I6060-01	MODTANK-102017-001	WATER	2	1000	100	3.1051	3.1051	3.02	3.1054	3.1054	0.0003	0.3



QC Batch# LB91006 Test: TPH Analysis Date: 10/23/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2288
pH Paper 0-14	W2248
Sodium Sulfate	EP1786
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP58006
LCSWD	10.00 ML	WP58005
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 71 Dessicator Time In : 12:25
1.0000 gram Balance: 1.0003 In Time: 11:45
Bal Check Time: 10:40

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 70 Dessicator Time Out: 14:00
1.0000 gram Balance: 1.0005 Out Time: 13:25
Bal Check Time: 14:30

1B 91006

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I6060

WorkList ID : 104536

Date : 10/23/2017 10:01:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/25/2017	Water	I5960-02 ✓	TPH	Conc H2SO4 4c	ARAM01	M11	GRAB	10/18/2017	1664A
10/26/2017	Water	I6020-01 ✓	TPH	Conc H2SO4 4c	ARAM01	M23	GRAB	10/19/2017	1664A
10/26/2017	Water	I6034-09 E	TPH	Conc H2SO4 4c	PSEG01	M52	BAKER-TANK	10/19/2017	1664A
10/25/2017	Water	I6060-01 F	TPH	Conc H2SO4 4c	ENTA05	O63	MODTANK-102017-001	10/20/2017	1664A

Date/Time 10/23/17 10:02 AM
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 10/23/17 11:40 AM
 Received by: [Signature]
 Relinquished by: [Signature]