


7471A

LB91019
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41925



A= 0.0000e+000

B= 2.4126e-004

C= -7.8590e-002

Rho= 0.9999686

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	387	0.000	387				
0.2	0.200	0.198	-0.002	1148	0.0 %	1148				
2.5	2.500	2.491	-0.009	10650	0.0 %	10650				
5.0	5.000	5.017	0.017	21121	0.0 %	21121				
7.5	7.500	7.443	-0.057	31177	0.0 %	31177	0			
10.0	10.000	10.036	0.036	41925	0.0 %	41925				

LB91019 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 23 Oct 2017 13:10:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	387	-	0.0000	7471A	PPB	23 Oct 2017 13:54:44	S	Std
0.2 - 1	502	1148	-	0.2000	7471A	PPB	23 Oct 2017 13:57:02	S	Std
2.5 - 1	52.5	10650	-	2.5000	7471A	PPB	23 Oct 2017 13:59:20	S	Std
5.0 - 1	55	21121	-	5.0000	7471A	PPB	23 Oct 2017 14:01:37	S	Std
7.5 - 1	57.5	31177	-	7.5000	7471A	PPB	23 Oct 2017 14:03:55	S	Std
10.0 - 1	510	41925	-	10.0000	7471A	PPB	23 Oct 2017 14:08:47	S	Std
ICV24 - 1	ICV24	17298	4.0947	-	7471A	PPB	23 Oct 2017 14:11:32	U	SMPL
ICB24 - 1	ICB24	191	-0.0325	-	7471A	PPB	23 Oct 2017 14:13:48	U	SMPL
CCV78 - 1	CCV78	22233	5.2852	-	7471A	PPB	23 Oct 2017 14:16:04	U	SMPL
CCB78 - 1	CCB78	209	-0.0282	-	7471A	PPB	23 Oct 2017 14:18:21	U	SMPL
CRA - 1	CRA	1131	0.1943	-	7471A	PPB	23 Oct 2017 14:20:37	U	SMPL
HighStd - 1	HighStd	42098	10.0778	-	7471A	PPB	23 Oct 2017 14:22:53	U	SMPL
ChkStd - 1	ChkStd	28216	6.7287	-	7471A	PPB	23 Oct 2017 14:25:09	U	SMPL
PB103516BL - 1	PBS	116	-0.0506	-	7471A	PPB	23 Oct 2017 14:27:28	U	SMPL
PB103516BS - 1	LCSS	15860	3.7477	-	7471A	PPB	23 Oct 2017 14:29:46	U	SMPL
I6019-01 - 1	SB-01	21804	5.1818	-	7471A	PPB	23 Oct 2017 14:32:03	U	SMPL
I6019-03 - 1	SB-02	61176	14.6805	-	7471A	PPB	23 Oct 2017 14:34:21	U	SMPL
I6019-08 - 1	SB-03	96	-0.0554	-	7471A	PPB	23 Oct 2017 14:36:40	U	SMPL
I6019-08DUP - 1	SB-03DUP	521	0.0471	-	7471A	PPB	23 Oct 2017 14:39:01	U	SMPL
I6019-08MS - 1	SB-03MS	15804	3.7342	-	7471A	PPB	23 Oct 2017 14:41:17	U	SMPL
CCV79 - 1	CCV79	21458	5.0983	-	7471A	PPB	23 Oct 2017 14:43:33	U	SMPL
CCB79 - 1	CCB79	162	-0.0395	-	7471A	PPB	23 Oct 2017 14:45:49	U	SMPL
I6019-08MSD - 1	SB-03MSD	14615	3.4474	-	7471A	PPB	23 Oct 2017 14:48:06	U	SMPL
I6025-01 - 1	SB-01	3205	0.6946	-	7471A	PPB	23 Oct 2017 14:50:22	U	SMPL
I6025-02 - 1	SB-02	3576	0.7841	-	7471A	PPB	23 Oct 2017 14:52:38	U	SMPL
I6025-03 - 1	SB-03	9151	2.1291	-	7471A	PPB	23 Oct 2017 14:54:54	U	SMPL
I6025-04 - 1	SB-04	2176	0.4464	-	7471A	PPB	23 Oct 2017 14:57:11	U	SMPL
I6025-05 - 1	SB-05	7135	1.6428	-	7471A	PPB	23 Oct 2017 14:59:28	U	SMPL
I6025-06 - 1	SB-06	10163	2.3733	-	7471A	PPB	23 Oct 2017 15:01:45	U	SMPL
I6028-01 - 1	15-40-RT4651	4394	0.9815	-	7471A	PPB	23 Oct 2017 15:04:02	U	SMPL
I6034-01 - 1	OILY-STONE	466	0.0338	-	7471A	PPB	23 Oct 2017 15:06:20	U	SMPL
I6044-01 - 1	NF-SOIL	2428	0.5072	-	7471A	PPB	23 Oct 2017 15:08:36	U	SMPL
CCV80 - 1	CCV80	21751	5.1690	-	7471A	PPB	23 Oct 2017 15:10:52	U	SMPL
CCB80 - 1	CCB80	172	-0.0371	-	7471A	PPB	23 Oct 2017 15:13:09	U	SMPL
PB103523BL - 1	PBS	333	0.0017	-	7471A	PPB	23 Oct 2017 15:15:26	U	SMPL
PB103523BS - 1	LCSS	15802	3.7337	-	7471A	PPB	23 Oct 2017 15:17:42	U	SMPL
I6029-01 - 1	EPE-108-2(60-62)	376	0.0121	-	7471A	PPB	23 Oct 2017 15:19:58	U	SMPL
I6029-02 - 1	EPE-108-2(65-67)	1127	0.1933	-	7471A	PPB	23 Oct 2017 15:22:15	U	SMPL
I6029-03 - 1	EPE-108-2(70-72)	845	0.1253	-	7471A	PPB	23 Oct 2017 15:24:31	U	SMPL
I6029-04 - 1	EPE-108-2(80-82)	572	0.0594	-	7471A	PPB	23 Oct 2017 15:26:48	U	SMPL
I6029-05 - 1	EPE-108-2(85-87)	1202	0.2114	-	7471A	PPB	23 Oct 2017 15:29:05	U	SMPL
I6029-06 - 1	EPE-P-5A(0-5)	8598	1.9957	-	7471A	PPB	23 Oct 2017 15:31:23	U	SMPL
I6029-07 - 1	EPE-P-5A(5-10)	1767	0.3477	-	7471A	PPB	23 Oct 2017 15:33:40	U	SMPL
I6029-07DUP - 1	EPE-P-5A(5-10)DUP	1572	0.3007	-	7471A	PPB	23 Oct 2017 15:35:57	U	SMPL
CCV81 - 1	CCV81	21682	5.1523	-	7471A	PPB	23 Oct 2017 15:38:13	U	SMPL
CCB81 - 1	CCB81	159	-0.0402	-	7471A	PPB	23 Oct 2017 15:40:29	U	SMPL
I6029-07MS - 1	EPE-P-5A(5-10)MS	17602	4.1680	-	7471A	PPB	23 Oct 2017 15:42:46	U	SMPL
I6029-07MSD - 1	EPE-P-5A(5-10)MSD	17431	4.1267	-	7471A	PPB	23 Oct 2017 15:45:02	U	SMPL
I6029-08 - 1	ENV-118-3(20-23)	439	0.0273	-	7471A	PPB	23 Oct 2017 15:47:19	U	SMPL
I6029-09 - 1	ENV-116-6(0-5)	6976	1.6044	-	7471A	PPB	23 Oct 2017 15:49:36	U	SMPL
I6029-10 - 1	ENV-116-6(5-10)	5035	1.1361	-	7471A	PPB	23 Oct 2017 15:51:52	U	SMPL
I6029-11 - 1	ENV-116-6(10-15)	1630	0.3147	-	7471A	PPB	23 Oct 2017 15:54:09	U	SMPL
I6029-12 - 1	ENV-116-6(15-20)	443	0.0283	-	7471A	PPB	23 Oct 2017 15:56:26	U	SMPL
I6029-13 - 1	ENV-116-6(20-30)	425	0.0239	-	7471A	PPB	23 Oct 2017 15:58:44	U	SMPL
I6029-14 - 1	ENV-116-6(30-38)	377	0.0124	-	7471A	PPB	23 Oct 2017 16:01:02	U	SMPL
I6029-15 - 1	ENV-118-3(0-5)	1939	0.3892	-	7471A	PPB	23 Oct 2017 16:03:18	U	SMPL
CCV82 - 1	CCV82	22280	5.2966	-	7471A	PPB	23 Oct 2017 16:05:34	U	SMPL
CCB82 - 1	CCB82	191	-0.0325	-	7471A	PPB	23 Oct 2017 16:07:51	U	SMPL
I6029-16 - 1	ENV-118-3(5-10)	633	0.0741	-	7471A	PPB	23 Oct 2017 16:10:08	U	SMPL
I6029-17 - 1	ENV-118-3(10-15)	488	0.0391	-	7471A	PPB	23 Oct 2017 16:12:25	U	SMPL
I6029-18 - 1	ENV-118-3(15-20)	619	0.0707	-	7471A	PPB	23 Oct 2017 16:14:41	U	SMPL
I6019-08LX5 - 1		373	0.0114	-	7471A	PPB	23 Oct 2017 16:16:57	U	SMPL
I6019-08A - 1		16958	4.0126	-	7471A	PPB	23 Oct 2017 16:19:14	U	SMPL
I6029-07LX5 - 1		526	0.0483	-	7471A	PPB	23 Oct 2017 16:21:31	U	SMPL
I6029-07A - 1		18688	4.4300	-	7471A	PPB	23 Oct 2017 16:23:48	U	SMPL
I6019-03DLX5 - 1		11958	2.8063	-	7471A	PPB	23 Oct 2017 16:26:06	U	SMPL
CCV83 - 1	CCV83	21843	5.1912	-	7471A	PPB	23 Oct 2017 16:28:24	U	SMPL
CCB83 - 1	CCB83	-39	-0.0880	-	7471A	PPB	23 Oct 2017 16:30:41	U	SMPL