



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91028

Temp.: 22.2 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/23/2017	17:15
2	CAL2	1	Water			7.00	10/23/2017	17:16
3	CAL3	1	Water			10.03	10/23/2017	17:19
4	ICV	1	Water			7.01	10/23/2017	17:22
5	CCV1	1	Water			2.01	10/23/2017	17:24
6	I6085-07	1	Solid	20.03	20	7.66	10/23/2017	17:25
7	I6091-03	1	Solid	20.04	20	7.64	10/23/2017	17:26
8	I6091-03DUP	1	Solid	20.03	20	7.65	10/23/2017	17:27
9	CCV2	1	Water			12.02	10/23/2017	17:29

2B91028

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6090

WorkList ID : 104554

Date : 10/23/2017 5:04:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/26/2017	Solid	I6085-07	Corrosivity	Cool 4 deg C	PSEG01	D11	ROLLOFF-28-20	10/23/2017	9045C
10/26/2017	Solid	I6091-03	Corrosivity	Cool 4 deg C	PSEG05	D52	HD-01-102317	10/23/2017	9045C

Date/Time 10-23-17 5:00 PM

Received by: JDRelinquished by: JD

Date/Time 10-23-17 2:30 PM

Received by: JDRelinquished by: JD