



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91030

Temp.: 20.1 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/23/2017	17:30
2	CAL2	1	Water			6.99	10/23/2017	17:31
3	CAL3	1	Water			10.02	10/23/2017	17:32
4	ICV	1	Water			7.00	10/23/2017	17:33
5	CCV1	1	Water			2.01	10/23/2017	17:34
6	I6090-01	1	Water			7.13	10/23/2017	17:35
7	I6090-01DUP	1	Water			7.14	10/23/2017	17:36
8	CCV2	1	Water			12.01	10/23/2017	17:37

1391030

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6080

WorkList ID : 104557

Date : 10/23/2017 5:05:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/23/2017	Water	I6090-01 -17	pH	Cool 4 deg C	PSEG01	D52	FADW29217L	10/23/2017	SM 4500-PH B

Date/Time 10-23-17 5:00 PM
Received by: 74 CP
Relinquished by:

Date/Time 10-23-17 6:35 PM
Received by: CP
Relinquished by: JA