


7470A

LB 91037
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34091



A= 0.0000e+000

B= 2.9428e-004

C= -1.6166e-002

Rho= 0.9999832

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	21	0.000	21				
0.2	0.200	0.203	0.003	746	0.0 %	746				
2.5	2.500	2.506	0.006	8570	0.0 %	8570				
5.0	5.000	5.026	0.026	17133	0.0 %	17133				
7.5	7.500	7.459	-0.041	25402	0.0 %	25402				
10.0	10.000	10.016	0.016	34091	0.0 %	34091				

LB91037 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Oct 2017 09:38:24

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.00007470A	PPB		24 Oct 2017 09:56:56	S	Std
0.2 - 1	50-2	746	-	0.20007470A	PPB		24 Oct 2017 09:59:05	S	Std
2.5 - 1	52.5	8570	-	2.50007470A	PPB		24 Oct 2017 10:01:13	S	Std
5.0 - 1	55	17133	-	5.00007470A	PPB		24 Oct 2017 10:03:21	S	Std
7.5 - 1	57.5	25402	-	7.50007470A	PPB		24 Oct 2017 10:05:29	S	Std
10.0 - 1	510	34091	-	10.00007470A	PPB		24 Oct 2017 10:07:38	S	Std
ICV25 - 1	ICV25	12505	3.6638	-7470A	PPB		24 Oct 2017 10:10:16	U	SMPL
ICB25 - 1	ICB25	-81	-0.0400	-7470A	PPB		24 Oct 2017 10:12:22	U	SMPL
CCV84 - 1	CCV84	17532	5.1431	-7470A	PPB		24 Oct 2017 10:14:29	U	SMPL
CCB84 - 1	CCB84	-174	-0.0674	-7470A	PPB		24 Oct 2017 10:16:36	U	SMPL
CRA - 1	CRA	717	0.1948	-7470A	PPB		24 Oct 2017 10:18:43	U	SMPL
HighStd - 1	HighStd	34593	10.1638	-7470A	PPB		24 Oct 2017 10:20:50	U	SMPL
ChkStd - 1	ChkStd	23921	7.0233	-7470A	PPB		24 Oct 2017 10:22:57	U	SMPL
PB103553BL - 1	PBW	-241	-0.0871	-7470A	PPB		24 Oct 2017 10:25:06	U	SMPL
PB103553BS - 1	LCSW	13306	3.8995	-7470A	PPB		24 Oct 2017 10:27:15	U	SMPL
PB103553TB - 1	PB103499TB	-134	-0.0556	-7470A	PPB		24 Oct 2017 10:29:23	U	SMPL
I6023-11 - 1	SB-01-COMP	-21	-0.0223	-7470A	PPB		24 Oct 2017 10:31:31	U	SMPL
I6023-12 - 1	SB-02-COMP	5	-0.0147	-7470A	PPB		24 Oct 2017 10:33:39	U	SMPL
I6023-13 - 1	SB-03-COMP	91	0.0106	-7470A	PPB		24 Oct 2017 10:35:47	U	SMPL
I6023-14 - 1	SB-04-COMP	46	-0.0026	-7470A	PPB		24 Oct 2017 10:37:54	U	SMPL
CCV85 - 1	CCV85	17850	5.2367	-7470A	PPB		24 Oct 2017 10:40:01	U	SMPL
CCB85 - 1	CCB85	-157	-0.0624	-7470A	PPB		24 Oct 2017 10:42:07	U	SMPL
I6023-15 - 1	SB-05-COMP	0	-0.0162	-7470A	PPB		24 Oct 2017 10:44:15	U	SMPL
I6024-17 - 1	SB-06-COMP	-6	-0.0179	-7470A	PPB		24 Oct 2017 10:46:22	U	SMPL
I6024-18 - 1	SB-07-COMP	-19	-0.0218	-7470A	PPB		24 Oct 2017 10:48:28	U	SMPL
I6024-19 - 1	SB-08-COMP	-7	-0.0182	-7470A	PPB		24 Oct 2017 10:50:35	U	SMPL
I6024-20 - 1	SB-09-COMP	-12	-0.0197	-7470A	PPB		24 Oct 2017 10:52:43	U	SMPL
I6024-21 - 1	SB-10-COMP	-8	-0.0185	-7470A	PPB		24 Oct 2017 10:54:50	U	SMPL
I6024-22 - 1	SB-11-COMP	-6	-0.0179	-7470A	PPB		24 Oct 2017 10:56:58	U	SMPL
I6024-23 - 1	SB-12-COMP	-11	-0.0194	-7470A	PPB		24 Oct 2017 10:59:06	U	SMPL
I6024-24 - 1	SB-13-COMP	-3	-0.0170	-7470A	PPB		24 Oct 2017 11:01:15	U	SMPL
I6028-04 - 1	15-40-RT4651	-6	-0.0179	-7470A	PPB		24 Oct 2017 11:03:22	U	SMPL
CCV86 - 1	CCV86	17499	5.1334	-7470A	PPB		24 Oct 2017 11:05:29	U	SMPL
CCB86 - 1	CCB86	-162	-0.0638	-7470A	PPB		24 Oct 2017 11:07:36	U	SMPL
I6034-03 - 1	OILY-STONE	-4	-0.0173	-7470A	PPB		24 Oct 2017 11:09:44	U	SMPL
I6034-05 - 1	OIL-DRUM	39	-0.0047	-7470A	PPB		24 Oct 2017 11:11:51	U	SMPL
I6034-10 - 1	BAKER-TANK	56	0.0003	-7470A	PPB		24 Oct 2017 11:13:58	U	SMPL
I6034-16 - 1	OILY-DEBRIS	62	0.0021	-7470A	PPB		24 Oct 2017 11:16:05	U	SMPL
I6044-06 - 1	NF-SOIL	95	0.0118	-7470A	PPB		24 Oct 2017 11:18:13	U	SMPL
I6044-06DUP - 1	NF-SOILDUP	91	0.0106	-7470A	PPB		24 Oct 2017 11:20:21	U	SMPL
I6044-06MS - 1	NF-SOILMS	12700	3.7212	-7470A	PPB		24 Oct 2017 11:22:29	U	SMPL
I6044-06MSD - 1	NF-SOILMSD	14258	4.1796	-7470A	PPB		24 Oct 2017 11:24:37	U	SMPL
PB103565BL - 1	PBW	15	-0.0118	-7470A	PPB		24 Oct 2017 11:26:44	U	SMPL
PB103565BS - 1	LCSW	13605	3.9875	-7470A	PPB		24 Oct 2017 11:28:54	U	SMPL
CCV87 - 1	CCV87	17689	5.1893	-7470A	PPB		24 Oct 2017 11:31:01	U	SMPL
CCB87 - 1	CCB87	-186	-0.0709	-7470A	PPB		24 Oct 2017 11:33:08	U	SMPL
PB103565TB - 1	PB103524TB	30	-0.0073	-7470A	PPB		24 Oct 2017 11:35:16	U	SMPL
I6060-01 - 1	MODTANK-102017-001	63	0.0024	-7470A	PPB		24 Oct 2017 11:37:23	U	SMPL
I6062-03 - 1	DRUM-1	72	0.0050	-7470A	PPB		24 Oct 2017 11:39:30	U	SMPL
I6063-01 - 1	S-1-A	62	0.0021	-7470A	PPB		24 Oct 2017 11:41:38	U	SMPL
I6064-03 - 1	RT-4248-COMP	4	-0.0150	-7470A	PPB		24 Oct 2017 11:43:45	U	SMPL
I6064-06 - 1	72-11989-COMP	-2	-0.0168	-7470A	PPB		24 Oct 2017 11:45:53	U	SMPL
I6064-09 - 1	72-11935-COMP	10	-0.0132	-7470A	PPB		24 Oct 2017 11:48:01	U	SMPL
I6065-01 - 1	DRUM	28	-0.0079	-7470A	PPB		24 Oct 2017 11:50:09	U	SMPL
I6069-03 - 1	OR-03-102017	75	0.0059	-7470A	PPB		24 Oct 2017 11:52:18	U	SMPL
I6070-03 - 1	OK-01-102017	17	-0.0112	-7470A	PPB		24 Oct 2017 11:54:25	U	SMPL
CCV88 - 1	CCV88	17739	5.2040	-7470A	PPB		24 Oct 2017 11:56:32	U	SMPL
CCB88 - 1	CCB88	-150	-0.0603	-7470A	PPB		24 Oct 2017 11:58:39	U	SMPL
I6070-03DUP - 1	OK-01-102017DUP	-11	-0.0194	-7470A	PPB		24 Oct 2017 12:00:48	U	SMPL
I6070-03MS - 1	OK-01-102017MS	12823	3.7574	-7470A	PPB		24 Oct 2017 12:02:55	U	SMPL
I6070-03MSD - 1	OK-01-102017MSD	13044	3.8224	-7470A	PPB		24 Oct 2017 12:05:03	U	SMPL
I6044-06LX5 - 1		7	-0.0141	-7470A	PPB		24 Oct 2017 12:07:11	U	SMPL
I6044-06A - 1		13944	4.0872	-7470A	PPB		24 Oct 2017 12:09:19	U	SMPL
I6070-03LX5 - 1		5	-0.0147	-7470A	PPB		24 Oct 2017 12:11:27	U	SMPL
I6070-03A - 1		13733	4.0251	-7470A	PPB		24 Oct 2017 12:13:35	U	SMPL
CCV89 - 1	CCV89	17780	5.2161	-7470A	PPB		24 Oct 2017 12:15:43	U	SMPL
CCB89 - 1	CCB89	-110	-0.0485	-7470A	PPB		24 Oct 2017 12:17:52	U	SMPL