


7470A

LB 91044
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34655



A= 0.0000e+000

B= 2.9032e-004

C= -3.6587e-002

Rho= 0.9999426

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	49	0.000	49				
0.2	0.200	0.193	-0.007	792	0.0 %	792				
2.5	2.500	2.537	0.037	8866	0.0 %	8866				
5.0	5.000	5.040	0.040	17486	0.0 %	17486				
7.5	7.500	7.427	-0.073	25710	0.0 %	25710				
10.0	10.000	10.024	0.024	34655	0.0 %	34655				

LB91044 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Oct 2017 13:21:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	49	-	0.0000	7470A	PPB	24 Oct 2017 13:59:28	S	Std
0.2 - 1	50.2	792	-	0.2000	7470A	PPB	24 Oct 2017 14:01:36	S	Std
2.5 - 1	52.5	8866	-	2.5000	7470A	PPB	24 Oct 2017 14:03:45	S	Std
5.0 - 1	55	17486	-	5.0000	7470A	PPB	24 Oct 2017 14:05:53	S	Std
7.5 - 1	57.5	25710	-	7.5000	7470A	PPB	24 Oct 2017 14:08:01	S	Std
10.0 - 1	510	34655	-	10.0000	7470A	PPB	24 Oct 2017 14:12:33	S	Std
ICV26 - 1	ICV26	13203	3.7965	-	7470A	PPB	24 Oct 2017 14:15:06	U	SMPL
ICB26 - 1	ICB26	-137	-0.0764	-	7470A	PPB	24 Oct 2017 14:17:13	U	SMPL
CCV90 - 1	CCV90	17230	4.9656	-	7470A	PPB	24 Oct 2017 14:19:20	U	SMPL
CCB90 - 1	CCB90	-130	-0.0743	-	7470A	PPB	24 Oct 2017 14:21:27	U	SMPL
CRA - 1	CRA	754	0.1823	-	7470A	PPB	24 Oct 2017 14:23:35	U	SMPL
HighStd - 1	HighStd	35288	10.2081	-	7470A	PPB	24 Oct 2017 14:25:41	U	SMPL
ChkStd - 1	ChkStd	23958	6.9188	-	7470A	PPB	24 Oct 2017 14:27:48	U	SMPL
PB103571BL - 1	PBW	-230	-0.1034	-	7470A	PPB	24 Oct 2017 14:29:57	U	SMPL
PB103571BS - 1	LCSW	13412	3.8571	-	7470A	PPB	24 Oct 2017 14:32:07	U	SMPL
I6056-01 - 1	SW-1	1378	0.3635	-	7470A	PPB	24 Oct 2017 14:34:14	U	SMPL
I6056-01DUP - 1	SW-1DUP	1385	0.3655	-	7470A	PPB	24 Oct 2017 14:36:22	U	SMPL
I6056-01MS - 1	SW-1MS	16220	4.6723	-	7470A	PPB	24 Oct 2017 14:38:31	U	SMPL
I6056-01MSD - 1	SW-1MSD	16347	4.7092	-	7470A	PPB	24 Oct 2017 14:40:39	U	SMPL
I6068-03 - 1	MW-03	-163	-0.0839	-	7470A	PPB	24 Oct 2017 14:42:46	U	SMPL
CCV91 - 1	CCV91	17423	5.0216	-	7470A	PPB	24 Oct 2017 14:44:54	U	SMPL
CCB91 - 1	CCB91	-136	-0.0761	-	7470A	PPB	24 Oct 2017 14:47:01	U	SMPL
I6068-05 - 1	MW-02	23	-0.0299	-	7470A	PPB	24 Oct 2017 14:49:08	U	SMPL
I6068-06 - 1	MW-03	-22	-0.0430	-	7470A	PPB	24 Oct 2017 14:51:15	U	SMPL
I6068-07 - 1	MW-02	-5	-0.0380	-	7470A	PPB	24 Oct 2017 14:53:22	U	SMPL
PB103594BL - 1	PBW	-52	-0.0517	-	7470A	PPB	24 Oct 2017 14:55:29	U	SMPL
PB103594BS - 1	LCSW	13497	3.8818	-	7470A	PPB	24 Oct 2017 14:57:36	U	SMPL
PB103594TB - 1	PB103566TB	-124	-0.0726	-	7470A	PPB	24 Oct 2017 14:59:44	U	SMPL
I6008-04 - 1	LT-TS-SPLP-1	101	-0.0073	-	7470A	PPB	24 Oct 2017 15:01:52	U	SMPL
I6008-05 - 1	LT-TS-SPLP-2	113	-0.0038	-	7470A	PPB	24 Oct 2017 15:04:00	U	SMPL
I6008-06 - 1	LT-TS-SPLP-3	120	-0.0017	-	7470A	PPB	24 Oct 2017 15:06:09	U	SMPL
I6008-06DUP - 1	LT-TS-SPLP-3DUP	123	-0.0009	-	7470A	PPB	24 Oct 2017 15:08:16	U	SMPL
CCV92 - 1	CCV92	16992	4.8965	-	7470A	PPB	24 Oct 2017 15:10:23	U	SMPL
CCB92 - 1	CCB92	-153	-0.0810	-	7470A	PPB	24 Oct 2017 15:12:30	U	SMPL
I6008-06MS - 1	LT-TS-SPLP-3MS	12916	3.7131	-	7470A	PPB	24 Oct 2017 15:14:38	U	SMPL
I6008-06MSD - 1	LT-TS-SPLP-3MSD	12486	3.5883	-	7470A	PPB	24 Oct 2017 15:16:45	U	SMPL
I6056-01LX5 - 1		152	0.0075	-	7470A	PPB	24 Oct 2017 15:18:52	U	SMPL
I6056-01A - 1		17200	4.9569	-	7470A	PPB	24 Oct 2017 15:21:00	U	SMPL
I6008-06LX5 - 1		-98	-0.0650	-	7470A	PPB	24 Oct 2017 15:23:08	U	SMPL
I6008-06A - 1		13615	3.9161	-	7470A	PPB	24 Oct 2017 15:25:16	U	SMPL
CCV93 - 1	CCV93	16882	4.8645	-	7470A	PPB	24 Oct 2017 15:27:24	U	SMPL
CCB93 - 1	CCB93	-81	-0.0601	-	7470A	PPB	24 Oct 2017 15:29:33	U	SMPL