



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91058

Temp.: 21.7 °C

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/24/2017	17:00
2	CAL2	1	Water			7.00	10/24/2017	17:01
3	CAL3	1	Water			10.04	10/24/2017	17:03
4	ICV	1	Water			6.99	10/24/2017	17:05
5	CCV1	1	Water			2.01	10/24/2017	17:06
6	I6111-02	1	Water			7.90	10/24/2017	17:09
7	I6111-02DUP	1	Water			7.91	10/24/2017	17:11
8	CCV2	1	Water			12.01	10/24/2017	17:13

WORKLIST(Hardcopy Internal Chain)

2B61058

WorkList Name : CORROSIONITY I6118 WorkList ID : 104614 Date : 10/24/2017 5:49:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2017	Water	I6111-02 -17	Corrosivity	Cool 4 deg C	PSEG01	D52	MARION26KV-3WATER	10/24/2017	9040C

Date/Time 10-24-17 5:00 PM
 Received by: JD
 Relinquished by: [Signature]

Date/Time 10-24-17 7:00 PM
 Received by: [Signature]
 Relinquished by: JP