



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91059

Temp.: 21.7 °C

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/24/2017	17:00
2	CAL2	1	Water			7.00	10/24/2017	17:01
3	CAL3	1	Water			10.04	10/24/2017	17:03
4	ICV	1	Water			6.99	10/24/2017	17:05
5	CCV1	1	Water			2.01	10/24/2017	17:06
6	I6111-01	1	Water			7.90	10/24/2017	17:09
7	I6111-01DUP	1	Water			7.91	10/24/2017	17:11
8	I6115-01	1	Water			7.41	10/24/2017	17:16
9	I6118-01	1	Water			7.56	10/24/2017	17:19
10	CCV2	1	Water			12.01	10/24/2017	17:20

WORKLIST(Hardcopy Internal Chain)

1391059

WorkList Name : PH I6109

WorkList ID : 104615

Date : 10/24/2017 5:50:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2017	Water	I6111-01 -A-	pH	Cool 4 deg C	PSEG01	D52	MARION26KV-3WATER	10/24/2017	SM 4500-PH B
10/24/2017	Water	I6115-01 -A-	pH	Cool 4 deg C	UNIO03	D11	STORM-DRAM	10/24/2017	SM 4500-PH B
10/24/2017	Water	I6118-01 -A-	pH	Cool 4 deg C	PSEG01	D52	FRAC-TANK-102517	10/24/2017	SM 4500-PH B

Date/Time 10-24-17 5:00 PM

Received by: JP

Relinquished by: SP

Date/Time 10-24-17 7:00 PM

Received by: GP

Relinquished by: JP